

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q3716

Project: Adoni

Run No.: LB138034

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138034BSW							
Bromide	mg/L	10	10.3		103	1	90-110	11/25/2025
Chloride	mg/L	3	3.10		103	1	90-110	11/25/2025
Fluoride	mg/L	2	2.10		105	1	90-110	11/25/2025
Nitrite	mg/L	3	3.10		103	1	90-110	11/25/2025
Nitrate	mg/L	2.5	2.50		100	1	90-110	11/25/2025
Sulfate	mg/L	15	15.4		103	1	90-110	11/25/2025
Orthophosphate as P	mg/L	5	4.70		94	1	90-110	11/25/2025

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q3716

Project: Adoni

Run No.: LB138041

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138041BS							
COD	mg/L	50	48.2		96	1	90-110	11/25/2025

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q3716

Project: Adoni

Run No.: LB138043

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138043BS							
BOD5	mg/L	198	187		94	1	84.6-115.4	11/26/2025