

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3716

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4290	4000	107	90 - 110	P	11/26/2025	13:18	LB138061
	Manganese	1960	2000	98	90 - 110	P	11/26/2025	13:18	LB138061
	Sodium	19500	20000	98	90 - 110	P	11/26/2025	13:18	LB138061
	Zinc	2040	2000	102	90 - 110	P	11/26/2025	13:18	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3716

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	86.9	100	87	80 - 120	P	11/26/2025	13:22	LB138061
	Manganese	21.2	20.0	106	80 - 120	P	11/26/2025	13:22	LB138061
	Sodium	1820	2000	91	80 - 120	P	11/26/2025	13:22	LB138061
	Zinc	43.2	40.0	108	80 - 120	P	11/26/2025	13:22	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3716

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4960	5000	99	90 - 110	P	11/26/2025	14:18	LB138061
	Manganese	2340	2500	94	90 - 110	P	11/26/2025	14:18	LB138061
	Sodium	23500	25000	94	90 - 110	P	11/26/2025	14:18	LB138061
	Zinc	2520	2500	101	90 - 110	P	11/26/2025	14:18	LB138061
CCV02	Iron	5010	5000	100	90 - 110	P	11/26/2025	15:39	LB138061
	Manganese	2330	2500	93	90 - 110	P	11/26/2025	15:39	LB138061
	Sodium	23400	25000	93	90 - 110	P	11/26/2025	15:39	LB138061
	Zinc	2560	2500	103	90 - 110	P	11/26/2025	15:39	LB138061
CCV03	Iron	5050	5000	101	90 - 110	P	11/26/2025	16:29	LB138061
	Manganese	2390	2500	96	90 - 110	P	11/26/2025	16:29	LB138061
	Sodium	24500	25000	98	90 - 110	P	11/26/2025	16:29	LB138061
	Zinc	2530	2500	101	90 - 110	P	11/26/2025	16:29	LB138061
CCV04	Iron	5090	5000	102	90 - 110	P	11/26/2025	17:25	LB138061
	Manganese	2370	2500	95	90 - 110	P	11/26/2025	17:25	LB138061
	Sodium	24300	25000	97	90 - 110	P	11/26/2025	17:25	LB138061
	Zinc	2540	2500	102	90 - 110	P	11/26/2025	17:25	LB138061
CCV05	Iron	4950	5000	99	90 - 110	P	11/26/2025	18:16	LB138061
	Manganese	2400	2500	96	90 - 110	P	11/26/2025	18:16	LB138061
	Sodium	23700	25000	95	90 - 110	P	11/26/2025	18:16	LB138061
	Zinc	2510	2500	100	90 - 110	P	11/26/2025	18:16	LB138061
CCV06	Iron	4950	5000	99	90 - 110	P	11/26/2025	19:07	LB138061
	Manganese	2350	2500	94	90 - 110	P	11/26/2025	19:07	LB138061
	Sodium	23700	25000	95	90 - 110	P	11/26/2025	19:07	LB138061
	Zinc	2490	2500	99	90 - 110	P	11/26/2025	19:07	LB138061
CCV07	Iron	5010	5000	100	90 - 110	P	11/26/2025	19:57	LB138061
	Manganese	2360	2500	94	90 - 110	P	11/26/2025	19:57	LB138061
	Sodium	24000	25000	96	90 - 110	P	11/26/2025	19:57	LB138061
	Zinc	2490	2500	100	90 - 110	P	11/26/2025	19:57	LB138061
CCV08	Iron	5000	5000	100	90 - 110	P	11/26/2025	20:10	LB138061
	Manganese	2320	2500	93	90 - 110	P	11/26/2025	20:10	LB138061
	Sodium	23500	25000	94	90 - 110	P	11/26/2025	20:10	LB138061
	Zinc	2500	2500	100	90 - 110	P	11/26/2025	20:10	LB138061