



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3716

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	91.9	100	92	65 - 135	P	11/26/2025	13:31	LB138061
	Manganese	21.9	20.0	110	65 - 135	P	11/26/2025	13:31	LB138061
	Sodium	1940	2000	97	65 - 135	P	11/26/2025	13:31	LB138061
	Zinc	44.5	40.0	111	65 - 135	P	11/26/2025	13:31	LB138061