

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q3716</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3716-02</u>	client id: <u>MW3MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3716-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	5680		3950		1500	115		P
Manganese	ug/L	75 - 125	4690		4420		100	269		P
Sodium	ug/L	75 - 125	34000		31100		1500	196		P
Zinc	ug/L	75 - 125	135		30.2		100	105		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q3716</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3716-02</u>	client id: <u>MW3MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3716-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	5790		3950		1500	123		P
Manganese	ug/L	75 - 125	4560		4420		100	134		P
Sodium	ug/L	75 - 125	34100		31100		1500	200		P
Zinc	ug/L	75 - 125	134		30.2		100	104		P