

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q3716

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170729BS							
Iron	ug/L	1500	1560		104	80 - 120	P
Manganese	ug/L	100	102		102	80 - 120	P
Sodium	ug/L	1500	1430		95	80 - 120	P
Zinc	ug/L	100	110		110	80 - 120	P