

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3716

Client ID : SSTDCCC040

Date Analyzed: 12/01/2025

Lab File ID: BP026197.D

Time Analyzed: 14:09

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	271172	7.66	1093510	10.43	683760	14.29
UPPER LIMIT	542344	8.16	2187020	10.925	1367520	14.789
LOWER LIMIT	135586	7.16	546755	9.925	341880	13.789
EPA SAMPLE NO.						
01 PB170743BL	220500	7.66	824685	10.43	520010	14.29
02 PB170743BS	273271	7.66	1112140	10.42	734282	14.30
03 PB170743BSD	243412	7.66	964799	10.43	610781	14.29
04 MW5	246581	7.66	985477	10.43	647616	14.28
05 MW3	216317	7.66	815646	10.42	502807	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.