

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3716
 Lab File ID: VN088370.D Date Analyzed: 12/01/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	197460	13.765				
UPPER LIMIT	394920	14.265				
LOWER LIMIT	98730	13.265				
EPA SAMPLE NO.						
MW5	195429	13.77				
MW3	234308	13.77				
MW7	206355	13.77				
VN1201WBL01	198799	13.77				
VN1201WBS01	207303	13.77				
VN1201WBSD01	196888	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.