

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3716
 Lab File ID: VN088396.D Date Analyzed: 12/02/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:00
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	206439	13.764				
UPPER LIMIT	412878	14.264				
LOWER LIMIT	103220	13.264				
EPA SAMPLE NO.						
MW5DL	241090	13.77				
MW4	184143	13.77				
MW7DL	196778	13.77				
VN1202WBL01	185686	13.77				
VN1202WBS02	188408	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.