

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3716

Lab File ID: VN088395.D

BFB Injection Date: 12/02/2025

Instrument ID: MSVOA_N

BFB Injection Time: 08:19

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.8
75	30.0 - 60.0% of mass 95	58.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.6
173	Less than 2.0% of mass 174	1.3 (1.8) 1
174	50.0 - 100.0% of mass 95	71.5
175	5.0 - 9.0% of mass 174	5.2 (7.3) 1
176	95.0 - 101.0% of mass 174	68.1 (95.3) 1
177	5.0 - 9.0% of mass 176	4.4 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN088396.D	12/02/2025	10:00
VN1202WBL01	VN1202WBL01	VN088397.D	12/02/2025	10:20
VN1202WBS02	VN1202WBS02	VN088400.D	12/02/2025	11:34
MW5DL	Q3716-01DL	VN088404.D	12/02/2025	12:55
MW7DL	Q3716-04DL	VN088405.D	12/02/2025	13:16
MW4	Q3716-03	VN088406.D	12/02/2025	13:36