

## **CASE NARRATIVE**

### **ATC Group Services LLC**

**Project Name: JHS**

**Project # N/A**

**Order ID # Q3717**

**Test Name: PCB Group1**

#### **A. Number of Samples and Date of Receipt:**

8 Solid samples were received on 11/24/2025.

#### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: PCB Group1. This data package contains results for PCB Group1.

#### **C. Analytical Techniques:**

The analyses were performed on instrument GCECD\_Q. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5  $\mu$ m df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25  $\mu$ m; Catalogue # 7HM-G017-11. The analysis of PCB Group1s was based on method 8082A and extraction was done based on method 3541.

#### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

Samples 1A1B1C, 2A2B2C, 3A3B3C, 4A4B4C, 5A5B5C, 6A6B6C, 7A7B7C and 8A8B8C are directly analyzed with dilution due to samples are Caulk matrix.

#### **E. Additional Comments:**

Less volume was taken for samples at the extraction due to CAULK matrix.

No MS-MSD performed as sample having CAULK matrix.

The temperature of the samples at the time of receipt was 15.2°C.

#### **F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_