

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ATCE02

Lab Code: ACE SDG NO.: Q3717

Instrument ID: ECD_Q Date(s) Analyzed: 11/18/2025 11/18/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.62	3.52	3.72	108487000
		2	3.70	3.60	3.80	79330000
		3	3.77	3.67	3.87	243616000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.76	3.66	3.86	197386000
		2	4.24	4.14	4.34	102466000
		3	4.52	4.42	4.62	207190000
		4	4.67	4.57	4.77	103833000
		5	4.81	4.71	4.91	84473000
Aroclor-1242	500	1	4.50	4.40	4.60	260736000
		2	4.52	4.42	4.62	380776000
		3	4.58	4.48	4.68	243244000
		4	4.67	4.57	4.77	194314000
		5	5.34	5.24	5.44	199781000
Aroclor-1248	500	1	4.50	4.40	4.60	190494000
		2	4.76	4.66	4.86	243288000
		3	4.95	4.85	5.05	293826000
		4	5.31	5.21	5.41	262878000
		5	5.34	5.24	5.44	338346000
Aroclor-1254	500	1	5.31	5.21	5.41	347360000
		2	5.53	5.43	5.63	541824000
		3	5.88	5.78	5.98	575830000
		4	6.16	6.06	6.26	422490000
		5	6.57	6.47	6.67	459632000
Aroclor-1262	500	1	6.87	6.77	6.97	466646000
		2	7.18	7.08	7.28	850324000
		3	7.45	7.35	7.55	560870000
		4	7.52	7.42	7.62	438268000
		5	8.02	7.92	8.12	283976000
Aroclor-1268	500	1	7.45	7.35	7.55	1020760000
		2	7.53	7.43	7.63	946750000
		3	7.71	7.61	7.81	777898000
		4	8.02	7.92	8.12	338330000
		5	8.31	8.21	8.41	2294740000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ATCE02

Lab Code: ACE SDG NO.: Q3717

Instrument ID: ECD_Q Date(s) Analyzed: 11/18/2025 11/18/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.06	2.96	3.16	114922000
		2	3.14	3.04	3.24	85299000
		3	3.21	3.11	3.31	258550000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.21	3.11	3.31	212032000
		2	3.89	3.79	3.99	216600000
		3	4.05	3.95	4.15	118883000
		4	4.14	4.04	4.24	93895400
		5	4.30	4.20	4.40	108074000
Aroclor-1242	500	1	3.87	3.77	3.97	289516000
		2	3.89	3.79	3.99	402834000
		3	4.05	3.95	4.15	233896000
		4	4.14	4.04	4.24	199829000
		5	4.64	4.54	4.74	263286000
Aroclor-1248	500	1	3.87	3.77	3.97	212876000
		2	4.10	4.00	4.20	284018000
		3	4.14	4.04	4.24	280112000
		4	4.30	4.20	4.40	346384000
		5	4.67	4.57	4.77	348638000
Aroclor-1254	500	1	4.64	4.54	4.74	535868000
		2	4.78	4.68	4.88	458782000
		3	5.17	5.07	5.27	759300000
		4	5.39	5.29	5.49	499818000
		5	5.80	5.70	5.90	692230000
Aroclor-1262	500	1	5.85	5.75	5.95	639510000
		2	6.09	5.99	6.19	594340000
		3	6.62	6.52	6.72	477658000
		4	6.68	6.58	6.78	894350000
		5	7.17	7.07	7.27	428332000
Aroclor-1268	500	1	6.62	6.52	6.72	1440040000
		2	6.68	6.58	6.78	1347710000
		3	6.88	6.78	6.98	1135760000
		4	7.17	7.07	7.27	507132000
		5	7.44	7.34	7.54	3541520000