

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3717
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071523.D **Time Analyzed:** 17:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	568.940	500.000	13.8
Aroclor-1016-2	4.515	4.416	4.616	577.180	500.000	15.4
Aroclor-1016-3	4.573	4.474	4.674	558.170	500.000	11.6
Aroclor-1016-4	4.663	4.564	4.764	570.920	500.000	14.2
Aroclor-1016-5	4.948	4.849	5.049	572.670	500.000	14.5
Aroclor-1260-1	6.042	5.943	6.143	537.760	500.000	7.6
Aroclor-1260-2	6.299	6.200	6.400	572.430	500.000	14.5
Aroclor-1260-3	6.651	6.552	6.752	564.980	500.000	13.0
Aroclor-1260-4	6.867	6.767	6.967	569.810	500.000	14.0
Aroclor-1260-5	7.174	7.075	7.275	578.590	500.000	15.7
Decachlorobiphenyl	8.536	8.437	8.637	49.730	50.000	-0.5
Tetrachloro-m-xylene	3.409	3.311	3.511	52.900	50.000	5.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3717
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071523.D **Time Analyzed:** 17:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.867	3.771	3.971	557.250	500.000	11.5
Aroclor-1016-2	3.883	3.788	3.988	578.030	500.000	15.6
Aroclor-1016-3	4.045	3.950	4.150	555.510	500.000	11.1
Aroclor-1016-4	4.093	3.998	4.198	548.220	500.000	9.6
Aroclor-1016-5	4.291	4.195	4.395	554.160	500.000	10.8
Aroclor-1260-1	5.291	5.197	5.397	550.760	500.000	10.2
Aroclor-1260-2	5.481	5.386	5.586	562.120	500.000	12.4
Aroclor-1260-3	5.623	5.530	5.730	549.070	500.000	9.8
Aroclor-1260-4	6.087	5.993	6.193	556.210	500.000	11.2
Aroclor-1260-5	6.332	6.238	6.438	556.510	500.000	11.3
Decachlorobiphenyl	7.651	7.559	7.759	47.250	50.000	-5.5
Tetrachloro-m-xylene	2.853	2.756	2.956	52.680	50.000	5.4