

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3718

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3970	4000	99	90 - 110	P	11/26/2025	13:18	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3718

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.6	12.0	97	80 - 120	P	11/26/2025	13:22	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3718

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4930	5000	98	90 - 110	P	11/26/2025	14:18	LB138061
CCV02	Lead	4980	5000	100	90 - 110	P	11/26/2025	15:39	LB138061
CCV03	Lead	4910	5000	98	90 - 110	P	11/26/2025	16:29	LB138061
CCV04	Lead	4930	5000	99	90 - 110	P	11/26/2025	17:25	LB138061
CCV05	Lead	4870	5000	97	90 - 110	P	11/26/2025	18:16	LB138061
CCV06	Lead	4900	5000	98	90 - 110	P	11/26/2025	19:07	LB138061
CCV07	Lead	4840	5000	97	90 - 110	P	11/26/2025	19:57	LB138061
CCV08	Lead	4890	5000	98	90 - 110	P	11/26/2025	20:10	LB138061