

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Total Quality Environmental Inc. **level:** low **sdg no.:** Q3718
contract: TOTA02 **lab code:** ACE
matrix: Water **sample id:** Q3716-02 **client id:** MW3MS
Percent Solids for Sample: NA **Spiked ID:** Q3716-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	474		1.41	J	500	94		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Total Quality Environmental Inc. **level:** low **sdg no.:** Q3718
contract: TOTA02 **lab code:** ACE
matrix: Water **sample id:** Q3716-02 **client id:** MW3MSD
Percent Solids for Sample: NA **Spiked ID:** Q3716-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	471		1.41	J	500	94		P