



# SAMPLE DATA

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-10A  
Lab Sample ID: Q3719-01

Date Collected: 11/24/25 14:40  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 76.2

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.19  | J    | 1  | 0.054 | 0.32       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:34 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-10B  
Lab Sample ID: Q3719-02

Date Collected: 11/24/25 15:00  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 86.5

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.12  | J    | 1  | 0.048 | 0.29       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:34 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-11A  
Lab Sample ID: Q3719-03

Date Collected: 11/24/25 14:05  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 80.3

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.12  | J    | 1  | 0.052 | 0.31       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:34 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-11B  
Lab Sample ID: Q3719-04

Date Collected: 11/24/25 14:15  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 89.2

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.091 | J    | 1  | 0.046 | 0.27       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:41 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-12A  
Lab Sample ID: Q3719-05

Date Collected: 11/24/25 12:20  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 67.8

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.44  |      | 1  | 0.060 | 0.35       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:41 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-12B  
Lab Sample ID: Q3719-06

Date Collected: 11/24/25 12:30  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 87.3

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.090 | J    | 1  | 0.048 | 0.28       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:41 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-14A  
Lab Sample ID: Q3719-07

Date Collected: 11/24/25 16:15  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 83.4

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.28  | J    | 1  | 0.049 | 0.29       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:41 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits



## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-14B  
Lab Sample ID: Q3719-08

Date Collected: 11/24/25 16:20  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 88.3

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.18  | J    | 1  | 0.047 | 0.28       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:41 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-15A  
Lab Sample ID: Q3719-09

Date Collected: 11/24/25 15:40  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 87.3

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.18  | J    | 1  | 0.046 | 0.28       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:49 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-15B  
Lab Sample ID: Q3719-10

Date Collected: 11/24/25 15:55  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 89.4

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.13  | J    | 1  | 0.046 | 0.27       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:49 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-16A  
Lab Sample ID: Q3719-11

Date Collected: 11/24/25 13:40  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 88

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.088 | J    | 1  | 0.046 | 0.28       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:49 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-16B  
Lab Sample ID: Q3719-12

Date Collected: 11/24/25 13:55  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 90.7

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.088 | J    | 1  | 0.045 | 0.27       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:49 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-17A  
Lab Sample ID: Q3719-13

Date Collected: 11/24/25 15:15  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 86.3

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.24  | J    | 1  | 0.047 | 0.28       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:49 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

Client: Remington & Vernick  
Project: Saddler Property  
Client Sample ID: SB-1125-17B  
Lab Sample ID: Q3719-14

Date Collected: 11/24/25 15:30  
Date Received: 11/25/25  
SDG No.: Q3719  
Matrix: SOIL  
% Solid: 84.3

| Parameter | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|-----------|-------|------|----|-------|------------|-------|----------------|----------------|----------|
| Cyanide   | 0.33  |      | 1  | 0.049 | 0.29       | mg/Kg | 11/26/25 08:15 | 11/26/25 12:49 | 9012B    |

Comments: \_\_\_\_\_

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits



# QC RESULT SUMMARY



## Initial and Continuing Calibration Verification

**Client:** Remington & Vernick

**SDG No.:** Q3719

**Project:** Saddler Property

**RunNo.:** LB138051

| Analyte                           | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|-----------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV1</b><br>Cyanide | mg/L  | 0.098  | 0.099      | 99         | 90-110                 | 11/26/2025    |
| Sample ID: <b>CCV1</b><br>Cyanide | mg/L  | 0.24   | 0.25       | 96         | 90-110                 | 11/26/2025    |
| Sample ID: <b>CCV2</b><br>Cyanide | mg/L  | 0.24   | 0.25       | 96         | 90-110                 | 11/26/2025    |
| Sample ID: <b>CCV3</b><br>Cyanide | mg/L  | 0.24   | 0.25       | 96         | 90-110                 | 11/26/2025    |
| Sample ID: <b>CCV4</b><br>Cyanide | mg/L  | 0.24   | 0.25       | 96         | 90-110                 | 11/26/2025    |

### Initial and Continuing Calibration Blank Summary

**Client:** Remington & Vernick

**SDG No.:** Q3719

**Project:** Saddler Property

**RunNo.:** LB138051

| Analyte                           | Units | Result | Acceptance Limits | Conc Qual | MDL     | RDL   | Analysis Date |
|-----------------------------------|-------|--------|-------------------|-----------|---------|-------|---------------|
| Sample ID: <b>ICB1</b><br>Cyanide | mg/L  | 0.0012 | 0.0025            | J         | 0.00096 | 0.005 | 11/26/2025    |
| Sample ID: <b>CCB1</b><br>Cyanide | mg/L  | 0.0016 | 0.0025            | J         | 0.00096 | 0.005 | 11/26/2025    |
| Sample ID: <b>CCB2</b><br>Cyanide | mg/L  | 0.0015 | 0.0025            | J         | 0.00096 | 0.005 | 11/26/2025    |
| Sample ID: <b>CCB3</b><br>Cyanide | mg/L  | 0.0015 | 0.0025            | J         | 0.00096 | 0.005 | 11/26/2025    |
| Sample ID: <b>CCB4</b><br>Cyanide | mg/L  | 0.0016 | 0.0025            | J         | 0.00096 | 0.005 | 11/26/2025    |

### Preparation Blank Summary

**Client:** Remington & Vernick

**SDG No.:** Q3719

**Project:** Saddler Property

| Analyte               | Units                      | Result | Acceptance<br>Limits | Conc<br>Qual | MDL   | RDL  | Analysis<br>Date |
|-----------------------|----------------------------|--------|----------------------|--------------|-------|------|------------------|
| Sample ID:<br>Cyanide | <b>PB170745BL</b><br>mg/Kg | 0.08   | 0.1250               | J            | 0.042 | 0.25 | 11/26/2025       |

## Matrix Spike Summary

|                   |                     |                                         |          |
|-------------------|---------------------|-----------------------------------------|----------|
| <b>Client:</b>    | Remington & Vernick | <b>SDG No.:</b>                         | Q3719    |
| <b>Project:</b>   | Saddler Property    | <b>Sample ID:</b>                       | Q3719-14 |
| <b>Client ID:</b> | SB-1125-17BMS       | <b>Percent Solids for Spike Sample:</b> | 84.3     |

| Analyte | Units | Acceptance<br>Limit %R | Spiked<br>Result | Conc.<br>Qualifier | Sample<br>Result | Conc.<br>Qualifier | Spike<br>Added | Dilution<br>Factor | %<br>Rec | Qual | Analysis<br>Date |
|---------|-------|------------------------|------------------|--------------------|------------------|--------------------|----------------|--------------------|----------|------|------------------|
| Cyanide | mg/Kg | 75-125                 | 2.40             |                    | 0.33             |                    | 2.3            | 1                  | 90       |      | 11/26/2025       |

## Matrix Spike Summary

|                   |                     |                                         |          |
|-------------------|---------------------|-----------------------------------------|----------|
| <b>Client:</b>    | Remington & Vernick | <b>SDG No.:</b>                         | Q3719    |
| <b>Project:</b>   | Saddler Property    | <b>Sample ID:</b>                       | Q3719-14 |
| <b>Client ID:</b> | SB-1125-17BMDS      | <b>Percent Solids for Spike Sample:</b> | 84.3     |

| Analyte | Units | Acceptance<br>Limit %R | Spiked<br>Result | Conc.<br>Qualifier | Sample<br>Result | Conc.<br>Qualifier | Spike<br>Added | Dilution<br>Factor | %<br>Rec | Qual | Analysis<br>Date |
|---------|-------|------------------------|------------------|--------------------|------------------|--------------------|----------------|--------------------|----------|------|------------------|
| Cyanide | mg/Kg | 75-125                 | 2.50             |                    | 0.33             |                    | 2.3            | 1                  | 94       |      | 11/26/2025       |

### Duplicate Sample Summary

|                   |                     |                                         |          |
|-------------------|---------------------|-----------------------------------------|----------|
| <b>Client:</b>    | Remington & Vernick | <b>SDG No.:</b>                         | Q3719    |
| <b>Project:</b>   | Saddler Property    | <b>Sample ID:</b>                       | Q3719-14 |
| <b>Client ID:</b> | SB-1125-17BDUP      | <b>Percent Solids for Spike Sample:</b> | 84.3     |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Cyanide | mg/Kg | +/-20               | 0.33             |                    | 0.33                |                    | 1                  | 0          |      | 11/26/2025       |

## Duplicate Sample Summary

|                   |                     |                                         |          |
|-------------------|---------------------|-----------------------------------------|----------|
| <b>Client:</b>    | Remington & Vernick | <b>SDG No.:</b>                         | Q3719    |
| <b>Project:</b>   | Saddler Property    | <b>Sample ID:</b>                       | Q3719-14 |
| <b>Client ID:</b> | SB-1125-17BMDS      | <b>Percent Solids for Spike Sample:</b> | 84.3     |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Cyanide | mg/Kg | +/-20               | 2.40             |                    | 2.50                |                    | 1                  | 4          |      | 11/26/2025       |

### Laboratory Control Sample Summary

**Client:** Remington & Vernick

**SDG No.:** Q3719

**Project:** Saddler Property

**Run No.:** LB138051

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | PB170745BS |            |        |                 |            |                 |                     |               |
| Cyanide   | mg/Kg      | 5          | 4.90   |                 | 98         | 1               | 85-115              | 11/26/2025    |



Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB138051**

|                  |                                                                |              |                       |
|------------------|----------------------------------------------------------------|--------------|-----------------------|
| Review By        | rubina                                                         | Review On    | 11/26/2025 3:46:10 PM |
| Supervise By     | Iwona                                                          | Supervise On | 11/26/2025 3:48:22 PM |
| SubDirectory     | LB138051                                                       | Test         | Cyanide               |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                               |              |                       |
| ICAL Standard    | WP115833,WP115834,WP115835,WP115836,WP115837,WP115838,WP115839 |              |                       |
| ICV Standard     | W3012                                                          |              |                       |
| CCV Standard     | WP115834                                                       |              |                       |
| ICSA Standard    | N/A                                                            |              |                       |
| CRI Standard     | N/A                                                            |              |                       |
| LCS Standard     | WP113838                                                       |              |                       |
| Chk Standard     | WP115157,WP114324,WP115840                                     |              |                       |

| Sr# | SampleId     | ClientID     | QcType | Date           | Comment | Operator | Status |
|-----|--------------|--------------|--------|----------------|---------|----------|--------|
| 1   | 0.0PPBCN     | 0.0PPBCN     | CAL1   | 11/26/25 10:20 |         | rubina   | OK     |
| 2   | 5.0PPBCN     | 5.0PPBCN     | CAL2   | 11/26/25 10:20 |         | rubina   | OK     |
| 3   | 10PPBCN      | 10PPBCN      | CAL3   | 11/26/25 10:20 |         | rubina   | OK     |
| 4   | 50PPBCN      | 50PPBCN      | CAL4   | 11/26/25 10:20 |         | rubina   | OK     |
| 5   | 100PPBCN     | 100PPBCN     | CAL5   | 11/26/25 10:20 |         | rubina   | OK     |
| 6   | 250PPBCN     | 250PPBCN     | CAL6   | 11/26/25 10:20 |         | rubina   | OK     |
| 7   | 500PPBCN     | 500PPBCN     | CAL7   | 11/26/25 10:20 |         | rubina   | OK     |
| 8   | ICV1         | ICV1         | ICV    | 11/26/25 12:26 |         | rubina   | OK     |
| 9   | ICB1         | ICB1         | ICB    | 11/26/25 12:26 |         | rubina   | OK     |
| 10  | CCV1         | CCV1         | CCV    | 11/26/25 12:26 |         | rubina   | OK     |
| 11  | CCB1         | CCB1         | CCB    | 11/26/25 12:26 |         | rubina   | OK     |
| 12  | PB170745BL   | PB170745BL   | MB     | 11/26/25 12:26 |         | rubina   | OK     |
| 13  | PB170745BS   | PB170745BS   | LCS    | 11/26/25 12:33 |         | rubina   | OK     |
| 14  | LOWPB170745  | LOWPB170745  | SAM    | 11/26/25 12:33 |         | rubina   | OK     |
| 15  | HIGHPB170745 | HIGHPB170745 | SAM    | 11/26/25 12:34 |         | rubina   | OK     |
| 16  | Q3719-01     | SB-1125-10A  | SAM    | 11/26/25 12:34 |         | rubina   | OK     |
| 17  | Q3719-02     | SB-1125-10B  | SAM    | 11/26/25 12:34 |         | rubina   | OK     |
| 18  | Q3719-03     | SB-1125-11A  | SAM    | 11/26/25 12:34 |         | rubina   | OK     |

Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB138051**

|                  |                                                                |              |                       |
|------------------|----------------------------------------------------------------|--------------|-----------------------|
| Review By        | rubina                                                         | Review On    | 11/26/2025 3:46:10 PM |
| Supervise By     | Iwona                                                          | Supervise On | 11/26/2025 3:48:22 PM |
| SubDirectory     | LB138051                                                       | Test         | Cyanide               |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                               |              |                       |
| ICAL Standard    | WP115833,WP115834,WP115835,WP115836,WP115837,WP115838,WP115839 |              |                       |
| ICV Standard     | W3012                                                          |              |                       |
| CCV Standard     | WP115834                                                       |              |                       |
| ICSA Standard    | N/A                                                            |              |                       |
| CRI Standard     | N/A                                                            |              |                       |
| LCS Standard     | WP113838                                                       |              |                       |
| Chk Standard     | WP115157,WP114324,WP115840                                     |              |                       |

|    |             |                |     |                |  |        |    |
|----|-------------|----------------|-----|----------------|--|--------|----|
| 19 | Q3719-05    | SB-1125-12A    | SAM | 11/26/25 12:41 |  | rubina | OK |
| 20 | Q3719-06    | SB-1125-12B    | SAM | 11/26/25 12:41 |  | rubina | OK |
| 21 | Q3719-04    | SB-1125-11B    | SAM | 11/26/25 12:41 |  | rubina | OK |
| 22 | CCV2        | CCV2           | CCV | 11/26/25 12:41 |  | rubina | OK |
| 23 | CCB2        | CCB2           | CCB | 11/26/25 12:41 |  | rubina | OK |
| 24 | Q3719-07    | SB-1125-14A    | SAM | 11/26/25 12:41 |  | rubina | OK |
| 25 | Q3719-08    | SB-1125-14B    | SAM | 11/26/25 12:41 |  | rubina | OK |
| 26 | Q3719-09    | SB-1125-15A    | SAM | 11/26/25 12:49 |  | rubina | OK |
| 27 | Q3719-10    | SB-1125-15B    | SAM | 11/26/25 12:49 |  | rubina | OK |
| 28 | Q3719-11    | SB-1125-16A    | SAM | 11/26/25 12:49 |  | rubina | OK |
| 29 | Q3719-12    | SB-1125-16B    | SAM | 11/26/25 12:49 |  | rubina | OK |
| 30 | Q3719-13    | SB-1125-17A    | SAM | 11/26/25 12:49 |  | rubina | OK |
| 31 | Q3719-14    | SB-1125-17B    | SAM | 11/26/25 12:49 |  | rubina | OK |
| 32 | Q3719-14DUP | SB-1125-17BDUP | DUP | 11/26/25 12:49 |  | rubina | OK |
| 33 | Q3719-14MS  | SB-1125-17BMS  | MS  | 11/26/25 12:49 |  | rubina | OK |
| 34 | CCV3        | CCV3           | CCV | 11/26/25 12:56 |  | rubina | OK |
| 35 | CCB3        | CCB3           | CCB | 11/26/25 12:56 |  | rubina | OK |
| 36 | Q3719-14MSD | SB-1125-17BMSD | MSD | 11/26/25 12:56 |  | rubina | OK |
| 37 | Q3721-01    | EO-2-112525    | SAM | 11/26/25 12:56 |  | rubina | OK |
| 38 | Q3723-01    | CC-1-112525    | SAM | 11/26/25 12:56 |  | rubina | OK |

Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB138051**

|                  |                                                                |              |                       |
|------------------|----------------------------------------------------------------|--------------|-----------------------|
| Review By        | rubina                                                         | Review On    | 11/26/2025 3:46:10 PM |
| Supervise By     | Iwona                                                          | Supervise On | 11/26/2025 3:48:22 PM |
| SubDirectory     | LB138051                                                       | Test         | Cyanide               |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                               |              |                       |
| ICAL Standard    | WP115833,WP115834,WP115835,WP115836,WP115837,WP115838,WP115839 |              |                       |
| ICV Standard     | W3012                                                          |              |                       |
| CCV Standard     | WP115834                                                       |              |                       |
| ICSA Standard    | N/A                                                            |              |                       |
| CRI Standard     | N/A                                                            |              |                       |
| LCS Standard     | WP113838                                                       |              |                       |
| Chk Standard     | WP115157,WP114324,WP115840                                     |              |                       |

|    |          |                  |     |                |                    |        |    |
|----|----------|------------------|-----|----------------|--------------------|--------|----|
| 39 | Q3725-01 | STOCKPILE-SAMPLE | SAM | 11/26/25 13:04 | %Solid is missing. | rubina | OK |
| 40 | CCV4     | CCV4             | CCV | 11/26/25 13:04 |                    | rubina | OK |
| 41 | CCB4     | CCB4             | CCB | 11/26/25 13:04 |                    | rubina | OK |

## LAB CHRONICLE

|                 |                     |                   |                        |
|-----------------|---------------------|-------------------|------------------------|
| <b>OrderID:</b> | Q3719               | <b>OrderDate:</b> | 11/25/2025 11:38:47 AM |
| <b>Client:</b>  | Remington & Vernick | <b>Project:</b>   | Saddler Property       |
| <b>Contact:</b> | Justin Zarzecki     | <b>Location:</b>  | D51                    |

| LabID           | ClientID           | Matrix      | Test    | Method | Sample Date               | Prep Date | Anal Date         | Received        |
|-----------------|--------------------|-------------|---------|--------|---------------------------|-----------|-------------------|-----------------|
| <b>Q3719-01</b> | <b>SB-1125-10A</b> | <b>SOIL</b> |         |        | <b>11/24/25<br/>14:40</b> |           |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B  |                           | 11/26/25  | 11/26/25<br>12:34 |                 |
| <b>Q3719-02</b> | <b>SB-1125-10B</b> | <b>SOIL</b> |         |        | <b>11/24/25<br/>15:00</b> |           |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B  |                           | 11/26/25  | 11/26/25<br>12:34 |                 |
| <b>Q3719-03</b> | <b>SB-1125-11A</b> | <b>SOIL</b> |         |        | <b>11/24/25<br/>14:05</b> |           |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B  |                           | 11/26/25  | 11/26/25<br>12:34 |                 |
| <b>Q3719-04</b> | <b>SB-1125-11B</b> | <b>SOIL</b> |         |        | <b>11/24/25<br/>14:15</b> |           |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B  |                           | 11/26/25  | 11/26/25<br>12:41 |                 |
| <b>Q3719-05</b> | <b>SB-1125-12A</b> | <b>SOIL</b> |         |        | <b>11/24/25<br/>12:20</b> |           |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B  |                           | 11/26/25  | 11/26/25<br>12:41 |                 |
| <b>Q3719-06</b> | <b>SB-1125-12B</b> | <b>SOIL</b> |         |        | <b>11/24/25<br/>12:30</b> |           |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B  |                           | 11/26/25  | 11/26/25<br>12:41 |                 |
| <b>Q3719-07</b> | <b>SB-1125-14A</b> | <b>SOIL</b> |         |        | <b>11/24/25<br/>16:15</b> |           |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B  |                           | 11/26/25  | 11/26/25<br>12:41 |                 |

## LAB CHRONICLE

|                 |                    |             |         |       |                                 |          |                   |                 |
|-----------------|--------------------|-------------|---------|-------|---------------------------------|----------|-------------------|-----------------|
| <b>Q3719-08</b> | <b>SB-1125-14B</b> | <b>SOIL</b> |         |       | <b>11/24/25</b><br><b>16:20</b> |          |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B |                                 | 11/26/25 | 11/26/25<br>12:41 |                 |
| <b>Q3719-09</b> | <b>SB-1125-15A</b> | <b>SOIL</b> |         |       | <b>11/24/25</b><br><b>15:40</b> |          |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B |                                 | 11/26/25 | 11/26/25<br>12:49 |                 |
| <b>Q3719-10</b> | <b>SB-1125-15B</b> | <b>SOIL</b> |         |       | <b>11/24/25</b><br><b>15:55</b> |          |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B |                                 | 11/26/25 | 11/26/25<br>12:49 |                 |
| <b>Q3719-11</b> | <b>SB-1125-16A</b> | <b>SOIL</b> |         |       | <b>11/24/25</b><br><b>13:40</b> |          |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B |                                 | 11/26/25 | 11/26/25<br>12:49 |                 |
| <b>Q3719-12</b> | <b>SB-1125-16B</b> | <b>SOIL</b> |         |       | <b>11/24/25</b><br><b>13:55</b> |          |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B |                                 | 11/26/25 | 11/26/25<br>12:49 |                 |
| <b>Q3719-13</b> | <b>SB-1125-17A</b> | <b>SOIL</b> |         |       | <b>11/24/25</b><br><b>15:15</b> |          |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B |                                 | 11/26/25 | 11/26/25<br>12:49 |                 |
| <b>Q3719-14</b> | <b>SB-1125-17B</b> | <b>SOIL</b> |         |       | <b>11/24/25</b><br><b>15:30</b> |          |                   | <b>11/25/25</b> |
|                 |                    |             | Cyanide | 9012B |                                 | 11/26/25 | 11/26/25<br>12:49 |                 |

SOP ID : M9012B-Total, Amenable and Reactive Cyanide-21

SDG No : N/A

Start Digest Date: 11/26/2025 Time : 08:15 Temp : 123 °C

Matrix : SOIL

End Digest Date: 11/26/2025 Time : 09:45 Temp : 127 °C

Pipette ID : WC

11 batch 11/26/2025 10:45 123 °C  
11/26/2025 11:45 127 °C

Balance ID : WC SC-7

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : MC-1,MC-2

Filter paper ID : N/A

Prep Technician Signature: *[Signature]*

Weigh By : JP

pH Meter ID : N/A

Supervisor Signature: *[Signature]*

| Standard Name     | MLS USED | STD REF. # FROM LOG |
|-------------------|----------|---------------------|
| LCSS              | 1.0ML    | WP113838            |
| MS/MSD SPIKE SOL. | 0.40ML   | WP113837            |
| PBS003            | 50.0ML   | W3112               |
| N/A               | N/A      | N/A                 |
| N/A               | N/A      | N/A                 |

| Chemical Used | ML/SAMPLE USED | Lot Number |
|---------------|----------------|------------|
| 0.25N NaOH    | 50.0ML         | WP113836   |
| 50% v/v H2SO4 | 5.0ML          | WP115334   |
| 51% w/v MgCL2 | 2.0ML          | WP115335   |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |

| LAB SAMPLE ID | CLIENT SAMPLE ID | Wt(g)/Vol(ml) | Comment  |
|---------------|------------------|---------------|----------|
| S0            | S0               | N/A           | N/A      |
| S5.0          | S5.0             | N/A           | N/A      |
| S10.0         | S10.0            | N/A           | N/A      |
| S100.0        | S100.0           | N/A           | N/A      |
| S250.0        | S250.0           | N/A           | N/A      |
| S500.0        | S500.0           | N/A           | N/A      |
| ICV           | ICV              | 0.5ML         | W3012    |
| ICB           | ICB              | N/A           | N/A      |
| CCV           | CCV              | N/A           | N/A      |
| CCB           | CCB              | N/A           | N/A      |
| Midrange      | Midrange         | N/A           | N/A      |
| HIGHSTD       | HIGHSTD          | 5.0ML         | WP113837 |
| LOWSTD        | LOWSTD           | 0.1ML         | WP113837 |

## Extraction Conformance/Non-Conformance Comments:

N/A

| Date / Time      | Prepped Sample Relinquished By/Location | Received By/Location |
|------------------|-----------------------------------------|----------------------|
| 11/26/2025 12:00 | JP / WC                                 | RM (WC)              |
|                  | Preparation Group                       | Analysis Group       |

| Lab Sample ID | Client Sample ID  | Initial Weight (g) | Final Vol (ml) | pH  | Sulfide | Oxidizing | Nitrate/ Nitrite | Comment | Prep Pos |
|---------------|-------------------|--------------------|----------------|-----|---------|-----------|------------------|---------|----------|
| PB170745BL    | PBS745            | 1.00               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| PB170745BS    | LCS745            | 1.00               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-01      | SB-1125-10A       | 1.03               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-02      | SB-1125-10B       | 1.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-03      | SB-1125-11A       | 1.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-04      | SB-1125-11B       | 1.03               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-05      | SB-1125-12A       | 1.04               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-06      | SB-1125-12B       | 1.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-07      | SB-1125-14A       | 1.02               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-08      | SB-1125-14B       | 1.02               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-09      | SB-1125-15A       | 1.04               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-10      | SB-1125-15B       | 1.02               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-11      | SB-1125-16A       | 1.03               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-12      | SB-1125-16B       | 1.04               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-13      | SB-1125-17A       | 1.03               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-14      | SB-1125-17B       | 1.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-14DUP   | SB-1125-17BDUP    | 1.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-14MS    | SB-1125-17BMS     | 1.03               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3719-14MSD   | SB-1125-17BMSD    | 1.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3721-01      | EO-2-112525       | 1.04               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3723-01      | CC-1-112525       | 1.02               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3725-01      | STOCKPILE-SAMPLES | 1.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q3725-04      | STOCKPILE-SAMPLES | 1.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |