

### Initial and Continuing Calibration Blank Summary

**Client:** Remington & Vernick

**SDG No.:** Q3719

**Project:** Saddler Property

**RunNo.:** LB138051

| Analyte                           | Units | Result | Acceptance<br>Limits | Conc<br>Qual | MDL     | RDL   | Analysis<br>Date |
|-----------------------------------|-------|--------|----------------------|--------------|---------|-------|------------------|
| Sample ID: <b>ICB1</b><br>Cyanide | mg/L  | 0.0012 | 0.0025               | J            | 0.00096 | 0.005 | 11/26/2025       |
| Sample ID: <b>CCB1</b><br>Cyanide | mg/L  | 0.0016 | 0.0025               | J            | 0.00096 | 0.005 | 11/26/2025       |
| Sample ID: <b>CCB2</b><br>Cyanide | mg/L  | 0.0015 | 0.0025               | J            | 0.00096 | 0.005 | 11/26/2025       |
| Sample ID: <b>CCB3</b><br>Cyanide | mg/L  | 0.0015 | 0.0025               | J            | 0.00096 | 0.005 | 11/26/2025       |
| Sample ID: <b>CCB4</b><br>Cyanide | mg/L  | 0.0016 | 0.0025               | J            | 0.00096 | 0.005 | 11/26/2025       |