



Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV130	Mercury	3.92	4.0	98	90 - 110	CV	11/26/2025	12:05	LB138045



Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV38	Mercury	5.16	5.0	103	90 - 110	CV	11/26/2025	12:13	LB138045
CCV39	Mercury	5.15	5.0	103	90 - 110	CV	11/26/2025	12:57	LB138045
CCV40	Mercury	5.09	5.0	102	90 - 110	CV	11/26/2025	13:27	LB138045
CCV41	Mercury	5.08	5.0	102	90 - 110	CV	11/26/2025	14:33	LB138045
CCV42	Mercury	5.12	5.0	102	90 - 110	CV	11/26/2025	14:47	LB138045

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7910	8000	99	90 - 110	P	11/26/2025	13:18	LB138061
	Antimony	4120	4000	103	90 - 110	P	11/26/2025	13:18	LB138061
	Arsenic	4050	4000	101	90 - 110	P	11/26/2025	13:18	LB138061
	Barium	7840	8000	98	90 - 110	P	11/26/2025	13:18	LB138061
	Beryllium	201	200	100	90 - 110	P	11/26/2025	13:18	LB138061
	Cadmium	2000	2000	100	90 - 110	P	11/26/2025	13:18	LB138061
	Calcium	19800	20000	99	90 - 110	P	11/26/2025	13:18	LB138061
	Chromium	869	800	109	90 - 110	P	11/26/2025	13:18	LB138061
	Cobalt	1990	2000	100	90 - 110	P	11/26/2025	13:18	LB138061
	Copper	1020	1000	102	90 - 110	P	11/26/2025	13:18	LB138061
	Iron	4290	4000	107	90 - 110	P	11/26/2025	13:18	LB138061
	Lead	3970	4000	99	90 - 110	P	11/26/2025	13:18	LB138061
	Magnesium	19600	20000	98	90 - 110	P	11/26/2025	13:18	LB138061
	Manganese	1960	2000	98	90 - 110	P	11/26/2025	13:18	LB138061
	Nickel	2030	2000	102	90 - 110	P	11/26/2025	13:18	LB138061
	Potassium	20000	20000	100	90 - 110	P	11/26/2025	13:18	LB138061
	Selenium	4120	4000	103	90 - 110	P	11/26/2025	13:18	LB138061
	Silver	1050	1000	105	90 - 110	P	11/26/2025	13:18	LB138061
	Sodium	19500	20000	98	90 - 110	P	11/26/2025	13:18	LB138061
	Thallium	3970	4000	99	90 - 110	P	11/26/2025	13:18	LB138061
	Vanadium	1960	2000	98	90 - 110	P	11/26/2025	13:18	LB138061
	Zinc	2040	2000	102	90 - 110	P	11/26/2025	13:18	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	99.6	100	100	80 - 120	P	11/26/2025	13:22	LB138061
	Antimony	52.7	50.0	106	80 - 120	P	11/26/2025	13:22	LB138061
	Arsenic	18.2	20.0	91	80 - 120	P	11/26/2025	13:22	LB138061
	Barium	98.9	100	99	80 - 120	P	11/26/2025	13:22	LB138061
	Beryllium	6.37	6.0	106	80 - 120	P	11/26/2025	13:22	LB138061
	Cadmium	6.05	6.0	101	80 - 120	P	11/26/2025	13:22	LB138061
	Calcium	2050	2000	102	80 - 120	P	11/26/2025	13:22	LB138061
	Chromium	10.5	10.0	105	80 - 120	P	11/26/2025	13:22	LB138061
	Cobalt	30.4	30.0	101	80 - 120	P	11/26/2025	13:22	LB138061
	Copper	22.3	20.0	112	80 - 120	P	11/26/2025	13:22	LB138061
	Iron	86.9	100	87	80 - 120	P	11/26/2025	13:22	LB138061
	Lead	11.6	12.0	97	80 - 120	P	11/26/2025	13:22	LB138061
	Magnesium	2160	2000	108	80 - 120	P	11/26/2025	13:22	LB138061
	Manganese	21.2	20.0	106	80 - 120	P	11/26/2025	13:22	LB138061
	Nickel	41.5	40.0	104	80 - 120	P	11/26/2025	13:22	LB138061
	Potassium	1920	2000	96	80 - 120	P	11/26/2025	13:22	LB138061
	Selenium	23.2	20.0	116	80 - 120	P	11/26/2025	13:22	LB138061
	Silver	10.7	10.0	107	80 - 120	P	11/26/2025	13:22	LB138061
	Sodium	1820	2000	91	80 - 120	P	11/26/2025	13:22	LB138061
	Thallium	39.0	40.0	98	80 - 120	P	11/26/2025	13:22	LB138061
	Vanadium	37.7	40.0	94	80 - 120	P	11/26/2025	13:22	LB138061
	Zinc	43.2	40.0	108	80 - 120	P	11/26/2025	13:22	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9780	10000	98	90 - 110	P	11/26/2025	14:18	LB138061
	Antimony	5000	5000	100	90 - 110	P	11/26/2025	14:18	LB138061
	Arsenic	4980	5000	100	90 - 110	P	11/26/2025	14:18	LB138061
	Barium	9390	10000	94	90 - 110	P	11/26/2025	14:18	LB138061
	Beryllium	240	250	96	90 - 110	P	11/26/2025	14:18	LB138061
	Cadmium	2450	2500	98	90 - 110	P	11/26/2025	14:18	LB138061
	Calcium	24200	25000	97	90 - 110	P	11/26/2025	14:18	LB138061
	Chromium	971	1000	97	90 - 110	P	11/26/2025	14:18	LB138061
	Cobalt	2440	2500	98	90 - 110	P	11/26/2025	14:18	LB138061
	Copper	1250	1250	100	90 - 110	P	11/26/2025	14:18	LB138061
	Iron	4960	5000	99	90 - 110	P	11/26/2025	14:18	LB138061
	Lead	4930	5000	98	90 - 110	P	11/26/2025	14:18	LB138061
	Magnesium	24100	25000	97	90 - 110	P	11/26/2025	14:18	LB138061
	Manganese	2340	2500	94	90 - 110	P	11/26/2025	14:18	LB138061
	Nickel	2440	2500	98	90 - 110	P	11/26/2025	14:18	LB138061
	Potassium	24600	25000	98	90 - 110	P	11/26/2025	14:18	LB138061
	Selenium	5070	5000	101	90 - 110	P	11/26/2025	14:18	LB138061
	Silver	1250	1250	100	90 - 110	P	11/26/2025	14:18	LB138061
	Sodium	23500	25000	94	90 - 110	P	11/26/2025	14:18	LB138061
	Thallium	5340	5000	107	90 - 110	P	11/26/2025	14:18	LB138061
Vanadium	2420	2500	97	90 - 110	P	11/26/2025	14:18	LB138061	
Zinc	2520	2500	101	90 - 110	P	11/26/2025	14:18	LB138061	
CCV02	Aluminum	9770	10000	98	90 - 110	P	11/26/2025	15:39	LB138061
	Antimony	5080	5000	102	90 - 110	P	11/26/2025	15:39	LB138061
	Arsenic	5070	5000	101	90 - 110	P	11/26/2025	15:39	LB138061
	Barium	9360	10000	94	90 - 110	P	11/26/2025	15:39	LB138061
	Beryllium	241	250	96	90 - 110	P	11/26/2025	15:39	LB138061
	Cadmium	2470	2500	99	90 - 110	P	11/26/2025	15:39	LB138061
	Calcium	24000	25000	96	90 - 110	P	11/26/2025	15:39	LB138061
	Chromium	967	1000	97	90 - 110	P	11/26/2025	15:39	LB138061
	Cobalt	2460	2500	99	90 - 110	P	11/26/2025	15:39	LB138061
	Copper	1260	1250	101	90 - 110	P	11/26/2025	15:39	LB138061
	Iron	5010	5000	100	90 - 110	P	11/26/2025	15:39	LB138061
	Lead	4980	5000	100	90 - 110	P	11/26/2025	15:39	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24100	25000	96	90 - 110	P	11/26/2025	15:39	LB138061
	Manganese	2330	2500	93	90 - 110	P	11/26/2025	15:39	LB138061
	Nickel	2460	2500	98	90 - 110	P	11/26/2025	15:39	LB138061
	Potassium	24800	25000	99	90 - 110	P	11/26/2025	15:39	LB138061
	Selenium	5160	5000	103	90 - 110	P	11/26/2025	15:39	LB138061
	Silver	1270	1250	101	90 - 110	P	11/26/2025	15:39	LB138061
	Sodium	23400	25000	93	90 - 110	P	11/26/2025	15:39	LB138061
	Thallium	5280	5000	106	90 - 110	P	11/26/2025	15:39	LB138061
	Vanadium	2410	2500	96	90 - 110	P	11/26/2025	15:39	LB138061
	Zinc	2560	2500	103	90 - 110	P	11/26/2025	15:39	LB138061
CCV03	Aluminum	9880	10000	99	90 - 110	P	11/26/2025	16:29	LB138061
	Antimony	5040	5000	101	90 - 110	P	11/26/2025	16:29	LB138061
	Arsenic	4990	5000	100	90 - 110	P	11/26/2025	16:29	LB138061
	Barium	9630	10000	96	90 - 110	P	11/26/2025	16:29	LB138061
	Beryllium	242	250	97	90 - 110	P	11/26/2025	16:29	LB138061
	Cadmium	2440	2500	97	90 - 110	P	11/26/2025	16:29	LB138061
	Calcium	24400	25000	97	90 - 110	P	11/26/2025	16:29	LB138061
	Chromium	967	1000	97	90 - 110	P	11/26/2025	16:29	LB138061
	Cobalt	2440	2500	97	90 - 110	P	11/26/2025	16:29	LB138061
	Copper	1250	1250	100	90 - 110	P	11/26/2025	16:29	LB138061
	Iron	5050	5000	101	90 - 110	P	11/26/2025	16:29	LB138061
	Lead	4910	5000	98	90 - 110	P	11/26/2025	16:29	LB138061
	Magnesium	24200	25000	97	90 - 110	P	11/26/2025	16:29	LB138061
	Manganese	2390	2500	96	90 - 110	P	11/26/2025	16:29	LB138061
	Nickel	2430	2500	97	90 - 110	P	11/26/2025	16:29	LB138061
	Potassium	25000	25000	100	90 - 110	P	11/26/2025	16:29	LB138061
	Selenium	5080	5000	102	90 - 110	P	11/26/2025	16:29	LB138061
	Silver	1250	1250	100	90 - 110	P	11/26/2025	16:29	LB138061
	Sodium	24500	25000	98	90 - 110	P	11/26/2025	16:29	LB138061
	Thallium	5120	5000	102	90 - 110	P	11/26/2025	16:29	LB138061
Vanadium	2440	2500	98	90 - 110	P	11/26/2025	16:29	LB138061	
Zinc	2530	2500	101	90 - 110	P	11/26/2025	16:29	LB138061	
CCV04	Aluminum	9820	10000	98	90 - 110	P	11/26/2025	17:25	LB138061
	Antimony	5060	5000	101	90 - 110	P	11/26/2025	17:25	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5010	5000	100	90 - 110	P	11/26/2025	17:25	LB138061
	Barium	9530	10000	95	90 - 110	P	11/26/2025	17:25	LB138061
	Beryllium	241	250	96	90 - 110	P	11/26/2025	17:25	LB138061
	Cadmium	2440	2500	98	90 - 110	P	11/26/2025	17:25	LB138061
	Calcium	24200	25000	97	90 - 110	P	11/26/2025	17:25	LB138061
	Chromium	973	1000	97	90 - 110	P	11/26/2025	17:25	LB138061
	Cobalt	2440	2500	98	90 - 110	P	11/26/2025	17:25	LB138061
	Copper	1250	1250	100	90 - 110	P	11/26/2025	17:25	LB138061
	Iron	5090	5000	102	90 - 110	P	11/26/2025	17:25	LB138061
	Lead	4930	5000	99	90 - 110	P	11/26/2025	17:25	LB138061
	Magnesium	24100	25000	96	90 - 110	P	11/26/2025	17:25	LB138061
	Manganese	2370	2500	95	90 - 110	P	11/26/2025	17:25	LB138061
	Nickel	2440	2500	98	90 - 110	P	11/26/2025	17:25	LB138061
	Potassium	25100	25000	100	90 - 110	P	11/26/2025	17:25	LB138061
	Selenium	5110	5000	102	90 - 110	P	11/26/2025	17:25	LB138061
	Silver	1270	1250	101	90 - 110	P	11/26/2025	17:25	LB138061
	Sodium	24300	25000	97	90 - 110	P	11/26/2025	17:25	LB138061
	Thallium	4940	5000	99	90 - 110	P	11/26/2025	17:25	LB138061
	Vanadium	2430	2500	97	90 - 110	P	11/26/2025	17:25	LB138061
Zinc	2540	2500	102	90 - 110	P	11/26/2025	17:25	LB138061	
CCV05	Aluminum	9950	10000	100	90 - 110	P	11/26/2025	18:16	LB138061
	Antimony	4990	5000	100	90 - 110	P	11/26/2025	18:16	LB138061
	Arsenic	4960	5000	99	90 - 110	P	11/26/2025	18:16	LB138061
	Barium	9630	10000	96	90 - 110	P	11/26/2025	18:16	LB138061
	Beryllium	243	250	97	90 - 110	P	11/26/2025	18:16	LB138061
	Cadmium	2420	2500	97	90 - 110	P	11/26/2025	18:16	LB138061
	Calcium	24600	25000	98	90 - 110	P	11/26/2025	18:16	LB138061
	Chromium	952	1000	95	90 - 110	P	11/26/2025	18:16	LB138061
	Cobalt	2420	2500	97	90 - 110	P	11/26/2025	18:16	LB138061
	Copper	1230	1250	99	90 - 110	P	11/26/2025	18:16	LB138061
	Iron	4950	5000	99	90 - 110	P	11/26/2025	18:16	LB138061
	Lead	4870	5000	97	90 - 110	P	11/26/2025	18:16	LB138061
	Magnesium	24500	25000	98	90 - 110	P	11/26/2025	18:16	LB138061
	Manganese	2400	2500	96	90 - 110	P	11/26/2025	18:16	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2410	2500	96	90 - 110	P	11/26/2025	18:16	LB138061
	Potassium	24400	25000	98	90 - 110	P	11/26/2025	18:16	LB138061
	Selenium	5040	5000	101	90 - 110	P	11/26/2025	18:16	LB138061
	Silver	1240	1250	99	90 - 110	P	11/26/2025	18:16	LB138061
	Sodium	23700	25000	95	90 - 110	P	11/26/2025	18:16	LB138061
	Thallium	4810	5000	96	90 - 110	P	11/26/2025	18:16	LB138061
	Vanadium	2460	2500	98	90 - 110	P	11/26/2025	18:16	LB138061
	Zinc	2510	2500	100	90 - 110	P	11/26/2025	18:16	LB138061
CCV06	Aluminum	9820	10000	98	90 - 110	P	11/26/2025	19:07	LB138061
	Antimony	5050	5000	101	90 - 110	P	11/26/2025	19:07	LB138061
	Arsenic	5020	5000	100	90 - 110	P	11/26/2025	19:07	LB138061
	Barium	9440	10000	94	90 - 110	P	11/26/2025	19:07	LB138061
	Beryllium	240	250	96	90 - 110	P	11/26/2025	19:07	LB138061
	Cadmium	2440	2500	98	90 - 110	P	11/26/2025	19:07	LB138061
	Calcium	24100	25000	96	90 - 110	P	11/26/2025	19:07	LB138061
	Chromium	946	1000	95	90 - 110	P	11/26/2025	19:07	LB138061
	Cobalt	2440	2500	97	90 - 110	P	11/26/2025	19:07	LB138061
	Copper	1250	1250	100	90 - 110	P	11/26/2025	19:07	LB138061
	Iron	4950	5000	99	90 - 110	P	11/26/2025	19:07	LB138061
	Lead	4900	5000	98	90 - 110	P	11/26/2025	19:07	LB138061
	Magnesium	24100	25000	96	90 - 110	P	11/26/2025	19:07	LB138061
	Manganese	2350	2500	94	90 - 110	P	11/26/2025	19:07	LB138061
	Nickel	2430	2500	97	90 - 110	P	11/26/2025	19:07	LB138061
	Potassium	24300	25000	97	90 - 110	P	11/26/2025	19:07	LB138061
	Selenium	5140	5000	103	90 - 110	P	11/26/2025	19:07	LB138061
	Silver	1240	1250	99	90 - 110	P	11/26/2025	19:07	LB138061
	Sodium	23700	25000	95	90 - 110	P	11/26/2025	19:07	LB138061
	Thallium	5330	5000	107	90 - 110	P	11/26/2025	19:07	LB138061
Vanadium	2420	2500	97	90 - 110	P	11/26/2025	19:07	LB138061	
Zinc	2490	2500	99	90 - 110	P	11/26/2025	19:07	LB138061	
CCV07	Aluminum	9660	10000	97	90 - 110	P	11/26/2025	19:57	LB138061
	Antimony	4980	5000	100	90 - 110	P	11/26/2025	19:57	LB138061
	Arsenic	4960	5000	99	90 - 110	P	11/26/2025	19:57	LB138061
	Barium	9510	10000	95	90 - 110	P	11/26/2025	19:57	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	239	250	96	90 - 110	P	11/26/2025	19:57	LB138061
	Cadmium	2390	2500	96	90 - 110	P	11/26/2025	19:57	LB138061
	Calcium	23800	25000	95	90 - 110	P	11/26/2025	19:57	LB138061
	Chromium	923	1000	92	90 - 110	P	11/26/2025	19:57	LB138061
	Cobalt	2400	2500	96	90 - 110	P	11/26/2025	19:57	LB138061
	Copper	1230	1250	98	90 - 110	P	11/26/2025	19:57	LB138061
	Iron	5010	5000	100	90 - 110	P	11/26/2025	19:57	LB138061
	Lead	4840	5000	97	90 - 110	P	11/26/2025	19:57	LB138061
	Magnesium	23600	25000	94	90 - 110	P	11/26/2025	19:57	LB138061
	Manganese	2360	2500	94	90 - 110	P	11/26/2025	19:57	LB138061
	Nickel	2390	2500	96	90 - 110	P	11/26/2025	19:57	LB138061
	Potassium	24600	25000	98	90 - 110	P	11/26/2025	19:57	LB138061
	Selenium	5040	5000	101	90 - 110	P	11/26/2025	19:57	LB138061
	Silver	1250	1250	100	90 - 110	P	11/26/2025	19:57	LB138061
	Sodium	24000	25000	96	90 - 110	P	11/26/2025	19:57	LB138061
	Thallium	4880	5000	98	90 - 110	P	11/26/2025	19:57	LB138061
Vanadium	2390	2500	96	90 - 110	P	11/26/2025	19:57	LB138061	
Zinc	2490	2500	100	90 - 110	P	11/26/2025	19:57	LB138061	
CCV08	Aluminum	9710	10000	97	90 - 110	P	11/26/2025	20:10	LB138061
	Antimony	5000	5000	100	90 - 110	P	11/26/2025	20:10	LB138061
	Arsenic	4970	5000	99	90 - 110	P	11/26/2025	20:10	LB138061
	Barium	9330	10000	93	90 - 110	P	11/26/2025	20:10	LB138061
	Beryllium	239	250	96	90 - 110	P	11/26/2025	20:10	LB138061
	Cadmium	2420	2500	97	90 - 110	P	11/26/2025	20:10	LB138061
	Calcium	23900	25000	96	90 - 110	P	11/26/2025	20:10	LB138061
	Chromium	918	1000	92	90 - 110	P	11/26/2025	20:10	LB138061
	Cobalt	2420	2500	97	90 - 110	P	11/26/2025	20:10	LB138061
	Copper	1240	1250	99	90 - 110	P	11/26/2025	20:10	LB138061
	Iron	5000	5000	100	90 - 110	P	11/26/2025	20:10	LB138061
	Lead	4890	5000	98	90 - 110	P	11/26/2025	20:10	LB138061
	Magnesium	23900	25000	96	90 - 110	P	11/26/2025	20:10	LB138061
	Manganese	2320	2500	93	90 - 110	P	11/26/2025	20:10	LB138061
	Nickel	2400	2500	96	90 - 110	P	11/26/2025	20:10	LB138061
	Potassium	24400	25000	98	90 - 110	P	11/26/2025	20:10	LB138061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5070	5000	101	90 - 110	P	11/26/2025	20:10	LB138061
	Silver	1250	1250	100	90 - 110	P	11/26/2025	20:10	LB138061
	Sodium	23500	25000	94	90 - 110	P	11/26/2025	20:10	LB138061
	Thallium	5020	5000	100	90 - 110	P	11/26/2025	20:10	LB138061
	Vanadium	2390	2500	96	90 - 110	P	11/26/2025	20:10	LB138061
	Zinc	2500	2500	100	90 - 110	P	11/26/2025	20:10	LB138061