



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	87	70 - 130	CV	11/26/2025	12:31	LB138045
CRI01	Aluminum	106	100	106	65 - 135	P	11/26/2025	13:31	LB138061
	Antimony	51.6	50.0	103	65 - 135	P	11/26/2025	13:31	LB138061
	Arsenic	20.1	20.0	101	65 - 135	P	11/26/2025	13:31	LB138061
	Barium	103	100	104	65 - 135	P	11/26/2025	13:31	LB138061
	Beryllium	6.48	6.0	108	65 - 135	P	11/26/2025	13:31	LB138061
	Cadmium	6.13	6.0	102	65 - 135	P	11/26/2025	13:31	LB138061
	Calcium	2120	2000	106	65 - 135	P	11/26/2025	13:31	LB138061
	Chromium	10.5	10.0	105	65 - 135	P	11/26/2025	13:31	LB138061
	Cobalt	31.0	30.0	104	65 - 135	P	11/26/2025	13:31	LB138061
	Copper	22.5	20.0	112	65 - 135	P	11/26/2025	13:31	LB138061
	Iron	91.9	100	92	65 - 135	P	11/26/2025	13:31	LB138061
	Lead	11.8	12.0	99	65 - 135	P	11/26/2025	13:31	LB138061
	Magnesium	2180	2000	109	65 - 135	P	11/26/2025	13:31	LB138061
	Manganese	21.9	20.0	110	65 - 135	P	11/26/2025	13:31	LB138061
	Nickel	42.4	40.0	106	65 - 135	P	11/26/2025	13:31	LB138061
	Potassium	2030	2000	101	65 - 135	P	11/26/2025	13:31	LB138061
	Selenium	22.3	20.0	112	65 - 135	P	11/26/2025	13:31	LB138061
	Silver	11.2	10.0	112	65 - 135	P	11/26/2025	13:31	LB138061
	Sodium	1940	2000	97	65 - 135	P	11/26/2025	13:31	LB138061
	Thallium	43.1	40.0	108	65 - 135	P	11/26/2025	13:31	LB138061
	Vanadium	40.1	40.0	100	65 - 135	P	11/26/2025	13:31	LB138061
	Zinc	44.5	40.0	111	65 - 135	P	11/26/2025	13:31	LB138061