



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB130	Mercury	0.076	+/-0.2	U	0.20	CV	11/26/2025	12:10	LB138045

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB38	Mercury	0.076	+/-0.2	U	0.20	CV	11/26/2025	12:18	LB138045
CCB39	Mercury	0.076	+/-0.2	U	0.20	CV	11/26/2025	13:02	LB138045
CCB40	Mercury	0.076	+/-0.2	U	0.20	CV	11/26/2025	13:32	LB138045
CCB41	Mercury	0.076	+/-0.2	U	0.20	CV	11/26/2025	14:38	LB138045
CCB42	Mercury	0.076	+/-0.2	U	0.20	CV	11/26/2025	14:54	LB138045

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	11/26/2025	13:26	LB138061
	Antimony	6.76	+/-25	U	50.0	P	11/26/2025	13:26	LB138061
	Arsenic	5.12	+/-10	U	20.0	P	11/26/2025	13:26	LB138061
	Barium	14.6	+/-50	U	100	P	11/26/2025	13:26	LB138061
	Beryllium	0.56	+/-3	U	6.00	P	11/26/2025	13:26	LB138061
	Cadmium	0.50	+/-3	U	6.00	P	11/26/2025	13:26	LB138061
	Calcium	234	+/-1000	U	2000	P	11/26/2025	13:26	LB138061
	Chromium	2.12	+/-5	U	10.0	P	11/26/2025	13:26	LB138061
	Cobalt	2.26	+/-15	U	30.0	P	11/26/2025	13:26	LB138061
	Copper	4.60	+/-10	U	20.0	P	11/26/2025	13:26	LB138061
	Iron	23.4	+/-50	U	100	P	11/26/2025	13:26	LB138061
	Lead	2.30	+/-6	U	12.0	P	11/26/2025	13:26	LB138061
	Magnesium	244	+/-1000	U	2000	P	11/26/2025	13:26	LB138061
	Manganese	5.94	+/-10	U	20.0	P	11/26/2025	13:26	LB138061
	Nickel	3.06	+/-20	U	40.0	P	11/26/2025	13:26	LB138061
	Potassium	918	+/-1000	U	2000	P	11/26/2025	13:26	LB138061
	Selenium	9.64	+/-10	U	20.0	P	11/26/2025	13:26	LB138061
	Silver	1.62	+/-5	U	10.0	P	11/26/2025	13:26	LB138061
	Sodium	868	+/-1000	U	2000	P	11/26/2025	13:26	LB138061
	Thallium	4.38	+/-20	U	40.0	P	11/26/2025	13:26	LB138061
	Vanadium	6.26	+/-20	U	40.0	P	11/26/2025	13:26	LB138061
	Zinc	16.7	+/-20	U	40.0	P	11/26/2025	13:26	LB138061

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	11/26/2025	14:22	LB138061
	Antimony	6.76	+/-25	U	50.0	P	11/26/2025	14:22	LB138061
	Arsenic	5.12	+/-10	U	20.0	P	11/26/2025	14:22	LB138061
	Barium	14.6	+/-50	U	100	P	11/26/2025	14:22	LB138061
	Beryllium	0.56	+/-3	U	6.00	P	11/26/2025	14:22	LB138061
	Cadmium	0.50	+/-3	U	6.00	P	11/26/2025	14:22	LB138061
	Calcium	234	+/-1000	U	2000	P	11/26/2025	14:22	LB138061
	Chromium	2.12	+/-5	U	10.0	P	11/26/2025	14:22	LB138061
	Cobalt	2.26	+/-15	U	30.0	P	11/26/2025	14:22	LB138061
	Copper	4.60	+/-10	U	20.0	P	11/26/2025	14:22	LB138061
	Iron	23.4	+/-50	U	100	P	11/26/2025	14:22	LB138061
	Lead	2.30	+/-6	U	12.0	P	11/26/2025	14:22	LB138061
	Magnesium	244	+/-1000	U	2000	P	11/26/2025	14:22	LB138061
	Manganese	5.94	+/-10	U	20.0	P	11/26/2025	14:22	LB138061
	Nickel	3.06	+/-20	U	40.0	P	11/26/2025	14:22	LB138061
	Potassium	918	+/-1000	U	2000	P	11/26/2025	14:22	LB138061
	Selenium	9.64	+/-10	U	20.0	P	11/26/2025	14:22	LB138061
	Silver	1.62	+/-5	U	10.0	P	11/26/2025	14:22	LB138061
	Sodium	868	+/-1000	U	2000	P	11/26/2025	14:22	LB138061
	Thallium	4.38	+/-20	U	40.0	P	11/26/2025	14:22	LB138061
	Vanadium	6.26	+/-20	U	40.0	P	11/26/2025	14:22	LB138061
	Zinc	16.7	+/-20	U	40.0	P	11/26/2025	14:22	LB138061
CCB02	Aluminum	11.3	+/-50	U	100	P	11/26/2025	15:43	LB138061
	Antimony	6.76	+/-25	U	50.0	P	11/26/2025	15:43	LB138061
	Arsenic	5.12	+/-10	U	20.0	P	11/26/2025	15:43	LB138061
	Barium	14.6	+/-50	U	100	P	11/26/2025	15:43	LB138061
	Beryllium	0.56	+/-3	U	6.00	P	11/26/2025	15:43	LB138061
	Cadmium	0.50	+/-3	U	6.00	P	11/26/2025	15:43	LB138061
	Calcium	234	+/-1000	U	2000	P	11/26/2025	15:43	LB138061
	Chromium	2.12	+/-5	U	10.0	P	11/26/2025	15:43	LB138061
	Cobalt	2.26	+/-15	U	30.0	P	11/26/2025	15:43	LB138061
	Copper	4.60	+/-10	U	20.0	P	11/26/2025	15:43	LB138061
	Iron	23.4	+/-50	U	100	P	11/26/2025	15:43	LB138061
	Lead	2.30	+/-6	U	12.0	P	11/26/2025	15:43	LB138061
	Magnesium	244	+/-1000	U	2000	P	11/26/2025	15:43	LB138061
	Manganese	5.94	+/-10	U	20.0	P	11/26/2025	15:43	LB138061
	Nickel	3.06	+/-20	U	40.0	P	11/26/2025	15:43	LB138061
	Potassium	918	+/-1000	U	2000	P	11/26/2025	15:43	LB138061
	Selenium	9.64	+/-10	U	20.0	P	11/26/2025	15:43	LB138061

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/26/2025	15:43	LB138061
	Sodium	868	+/-1000	U	2000	P	11/26/2025	15:43	LB138061
	Thallium	4.38	+/-20	U	40.0	P	11/26/2025	15:43	LB138061
	Vanadium	6.26	+/-20	U	40.0	P	11/26/2025	15:43	LB138061
	Zinc	16.7	+/-20	U	40.0	P	11/26/2025	15:43	LB138061
CCB03	Aluminum	11.3	+/-50	U	100	P	11/26/2025	16:33	LB138061
	Antimony	6.76	+/-25	U	50.0	P	11/26/2025	16:33	LB138061
	Arsenic	5.12	+/-10	U	20.0	P	11/26/2025	16:33	LB138061
	Barium	14.6	+/-50	U	100	P	11/26/2025	16:33	LB138061
	Beryllium	0.56	+/-3	U	6.00	P	11/26/2025	16:33	LB138061
	Cadmium	0.50	+/-3	U	6.00	P	11/26/2025	16:33	LB138061
	Calcium	234	+/-1000	U	2000	P	11/26/2025	16:33	LB138061
	Chromium	2.12	+/-5	U	10.0	P	11/26/2025	16:33	LB138061
	Cobalt	2.26	+/-15	U	30.0	P	11/26/2025	16:33	LB138061
	Copper	4.60	+/-10	U	20.0	P	11/26/2025	16:33	LB138061
	Iron	23.4	+/-50	U	100	P	11/26/2025	16:33	LB138061
	Lead	2.30	+/-6	U	12.0	P	11/26/2025	16:33	LB138061
	Magnesium	244	+/-1000	U	2000	P	11/26/2025	16:33	LB138061
	Manganese	5.94	+/-10	U	20.0	P	11/26/2025	16:33	LB138061
	Nickel	3.06	+/-20	U	40.0	P	11/26/2025	16:33	LB138061
	Potassium	918	+/-1000	U	2000	P	11/26/2025	16:33	LB138061
	Selenium	9.64	+/-10	U	20.0	P	11/26/2025	16:33	LB138061
	Silver	1.62	+/-5	U	10.0	P	11/26/2025	16:33	LB138061
	Sodium	868	+/-1000	U	2000	P	11/26/2025	16:33	LB138061
	Thallium	4.38	+/-20	U	40.0	P	11/26/2025	16:33	LB138061
	Vanadium	6.26	+/-20	U	40.0	P	11/26/2025	16:33	LB138061
	Zinc	16.7	+/-20	U	40.0	P	11/26/2025	16:33	LB138061
CCB04	Aluminum	11.3	+/-50	U	100	P	11/26/2025	17:29	LB138061
	Antimony	6.76	+/-25	U	50.0	P	11/26/2025	17:29	LB138061
	Arsenic	5.12	+/-10	U	20.0	P	11/26/2025	17:29	LB138061
	Barium	14.6	+/-50	U	100	P	11/26/2025	17:29	LB138061
	Beryllium	0.56	+/-3	U	6.00	P	11/26/2025	17:29	LB138061
	Cadmium	0.50	+/-3	U	6.00	P	11/26/2025	17:29	LB138061
	Calcium	234	+/-1000	U	2000	P	11/26/2025	17:29	LB138061
	Chromium	2.12	+/-5	U	10.0	P	11/26/2025	17:29	LB138061
	Cobalt	2.26	+/-15	U	30.0	P	11/26/2025	17:29	LB138061
	Copper	4.60	+/-10	U	20.0	P	11/26/2025	17:29	LB138061
	Iron	23.4	+/-50	U	100	P	11/26/2025	17:29	LB138061
	Lead	2.30	+/-6	U	12.0	P	11/26/2025	17:29	LB138061

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/26/2025	17:29	LB138061
	Manganese	5.94	+/-10	U	20.0	P	11/26/2025	17:29	LB138061
	Nickel	3.06	+/-20	U	40.0	P	11/26/2025	17:29	LB138061
	Potassium	918	+/-1000	U	2000	P	11/26/2025	17:29	LB138061
	Selenium	9.64	+/-10	U	20.0	P	11/26/2025	17:29	LB138061
	Silver	1.62	+/-5	U	10.0	P	11/26/2025	17:29	LB138061
	Sodium	868	+/-1000	U	2000	P	11/26/2025	17:29	LB138061
	Thallium	4.38	+/-20	U	40.0	P	11/26/2025	17:29	LB138061
	Vanadium	6.26	+/-20	U	40.0	P	11/26/2025	17:29	LB138061
	Zinc	16.7	+/-20	U	40.0	P	11/26/2025	17:29	LB138061
CCB05	Aluminum	11.3	+/-50	U	100	P	11/26/2025	18:20	LB138061
	Antimony	6.76	+/-25	U	50.0	P	11/26/2025	18:20	LB138061
	Arsenic	5.12	+/-10	U	20.0	P	11/26/2025	18:20	LB138061
	Barium	14.6	+/-50	U	100	P	11/26/2025	18:20	LB138061
	Beryllium	0.56	+/-3	U	6.00	P	11/26/2025	18:20	LB138061
	Cadmium	0.50	+/-3	U	6.00	P	11/26/2025	18:20	LB138061
	Calcium	234	+/-1000	U	2000	P	11/26/2025	18:20	LB138061
	Chromium	2.12	+/-5	U	10.0	P	11/26/2025	18:20	LB138061
	Cobalt	2.26	+/-15	U	30.0	P	11/26/2025	18:20	LB138061
	Copper	4.60	+/-10	U	20.0	P	11/26/2025	18:20	LB138061
	Iron	23.4	+/-50	U	100	P	11/26/2025	18:20	LB138061
	Lead	2.30	+/-6	U	12.0	P	11/26/2025	18:20	LB138061
	Magnesium	244	+/-1000	U	2000	P	11/26/2025	18:20	LB138061
	Manganese	5.94	+/-10	U	20.0	P	11/26/2025	18:20	LB138061
	Nickel	3.06	+/-20	U	40.0	P	11/26/2025	18:20	LB138061
	Potassium	918	+/-1000	U	2000	P	11/26/2025	18:20	LB138061
	Selenium	9.64	+/-10	U	20.0	P	11/26/2025	18:20	LB138061
	Silver	1.62	+/-5	U	10.0	P	11/26/2025	18:20	LB138061
	Sodium	868	+/-1000	U	2000	P	11/26/2025	18:20	LB138061
	Thallium	4.38	+/-20	U	40.0	P	11/26/2025	18:20	LB138061
	Vanadium	6.26	+/-20	U	40.0	P	11/26/2025	18:20	LB138061
	Zinc	16.7	+/-20	U	40.0	P	11/26/2025	18:20	LB138061
CCB06	Aluminum	11.3	+/-50	U	100	P	11/26/2025	19:11	LB138061
	Antimony	6.76	+/-25	U	50.0	P	11/26/2025	19:11	LB138061
	Arsenic	5.12	+/-10	U	20.0	P	11/26/2025	19:11	LB138061
	Barium	14.6	+/-50	U	100	P	11/26/2025	19:11	LB138061
	Beryllium	0.56	+/-3	U	6.00	P	11/26/2025	19:11	LB138061
	Cadmium	0.50	+/-3	U	6.00	P	11/26/2025	19:11	LB138061
	Calcium	234	+/-1000	U	2000	P	11/26/2025	19:11	LB138061

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	11/26/2025	19:11	LB138061
	Cobalt	2.26	+/-15	U	30.0	P	11/26/2025	19:11	LB138061
	Copper	4.60	+/-10	U	20.0	P	11/26/2025	19:11	LB138061
	Iron	23.4	+/-50	U	100	P	11/26/2025	19:11	LB138061
	Lead	2.30	+/-6	U	12.0	P	11/26/2025	19:11	LB138061
	Magnesium	244	+/-1000	U	2000	P	11/26/2025	19:11	LB138061
	Manganese	5.94	+/-10	U	20.0	P	11/26/2025	19:11	LB138061
	Nickel	3.06	+/-20	U	40.0	P	11/26/2025	19:11	LB138061
	Potassium	918	+/-1000	U	2000	P	11/26/2025	19:11	LB138061
	Selenium	9.64	+/-10	U	20.0	P	11/26/2025	19:11	LB138061
	Silver	1.62	+/-5	U	10.0	P	11/26/2025	19:11	LB138061
	Sodium	868	+/-1000	U	2000	P	11/26/2025	19:11	LB138061
	Thallium	4.38	+/-20	U	40.0	P	11/26/2025	19:11	LB138061
	Vanadium	6.26	+/-20	U	40.0	P	11/26/2025	19:11	LB138061
	Zinc	16.7	+/-20	U	40.0	P	11/26/2025	19:11	LB138061
CCB07	Aluminum	11.3	+/-50	U	100	P	11/26/2025	20:02	LB138061
	Antimony	6.76	+/-25	U	50.0	P	11/26/2025	20:02	LB138061
	Arsenic	5.12	+/-10	U	20.0	P	11/26/2025	20:02	LB138061
	Barium	14.6	+/-50	U	100	P	11/26/2025	20:02	LB138061
	Beryllium	0.56	+/-3	U	6.00	P	11/26/2025	20:02	LB138061
	Cadmium	0.50	+/-3	U	6.00	P	11/26/2025	20:02	LB138061
	Calcium	234	+/-1000	U	2000	P	11/26/2025	20:02	LB138061
	Chromium	2.12	+/-5	U	10.0	P	11/26/2025	20:02	LB138061
	Cobalt	2.26	+/-15	U	30.0	P	11/26/2025	20:02	LB138061
	Copper	4.60	+/-10	U	20.0	P	11/26/2025	20:02	LB138061
	Iron	23.4	+/-50	U	100	P	11/26/2025	20:02	LB138061
	Lead	2.30	+/-6	U	12.0	P	11/26/2025	20:02	LB138061
	Magnesium	244	+/-1000	U	2000	P	11/26/2025	20:02	LB138061
	Manganese	5.94	+/-10	U	20.0	P	11/26/2025	20:02	LB138061
	Nickel	3.06	+/-20	U	40.0	P	11/26/2025	20:02	LB138061
	Potassium	918	+/-1000	U	2000	P	11/26/2025	20:02	LB138061
	Selenium	9.64	+/-10	U	20.0	P	11/26/2025	20:02	LB138061
	Silver	1.62	+/-5	U	10.0	P	11/26/2025	20:02	LB138061
	Sodium	868	+/-1000	U	2000	P	11/26/2025	20:02	LB138061
	Thallium	4.38	+/-20	U	40.0	P	11/26/2025	20:02	LB138061
	Vanadium	6.26	+/-20	U	40.0	P	11/26/2025	20:02	LB138061
	Zinc	16.7	+/-20	U	40.0	P	11/26/2025	20:02	LB138061
CCB08	Aluminum	11.3	+/-50	U	100	P	11/26/2025	20:14	LB138061
	Antimony	6.76	+/-25	U	50.0	P	11/26/2025	20:14	LB138061

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	11/26/2025	20:14	LB138061
	Barium	14.6	+/-50	U	100	P	11/26/2025	20:14	LB138061
	Beryllium	0.56	+/-3	U	6.00	P	11/26/2025	20:14	LB138061
	Cadmium	0.50	+/-3	U	6.00	P	11/26/2025	20:14	LB138061
	Calcium	234	+/-1000	U	2000	P	11/26/2025	20:14	LB138061
	Chromium	2.12	+/-5	U	10.0	P	11/26/2025	20:14	LB138061
	Cobalt	2.26	+/-15	U	30.0	P	11/26/2025	20:14	LB138061
	Copper	4.60	+/-10	U	20.0	P	11/26/2025	20:14	LB138061
	Iron	23.4	+/-50	U	100	P	11/26/2025	20:14	LB138061
	Lead	2.30	+/-6	U	12.0	P	11/26/2025	20:14	LB138061
	Magnesium	244	+/-1000	U	2000	P	11/26/2025	20:14	LB138061
	Manganese	5.94	+/-10	U	20.0	P	11/26/2025	20:14	LB138061
	Nickel	3.06	+/-20	U	40.0	P	11/26/2025	20:14	LB138061
	Potassium	918	+/-1000	U	2000	P	11/26/2025	20:14	LB138061
	Selenium	9.64	+/-10	U	20.0	P	11/26/2025	20:14	LB138061
	Silver	1.62	+/-5	U	10.0	P	11/26/2025	20:14	LB138061
	Sodium	868	+/-1000	U	2000	P	11/26/2025	20:14	LB138061
	Thallium	4.38	+/-20	U	40.0	P	11/26/2025	20:14	LB138061
	Vanadium	6.26	+/-20	U	40.0	P	11/26/2025	20:14	LB138061
	Zinc	16.7	+/-20	U	40.0	P	11/26/2025	20:14	LB138061