

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170733BS Mercury	mg/Kg	0.28	0.30		105	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Remington & Vernick

SDG No.: Q3719

Contract: REMI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170744BS							
Aluminum	mg/Kg	100	102		102	80 - 120	P
Antimony	mg/Kg	40.0	42.3		106	80 - 120	P
Arsenic	mg/Kg	40.0	41.2		103	80 - 120	P
Barium	mg/Kg	10.0	9.51		95	80 - 120	P
Beryllium	mg/Kg	10.0	10.2		102	80 - 120	P
Cadmium	mg/Kg	10.0	10.1		101	80 - 120	P
Calcium	mg/Kg	50.0	53.3	J	107	80 - 120	P
Chromium	mg/Kg	20.0	20.1		100	80 - 120	P
Cobalt	mg/Kg	10.0	10.1		101	80 - 120	P
Copper	mg/Kg	15.0	16.0		107	80 - 120	P
Iron	mg/Kg	150	155		103	80 - 120	P
Lead	mg/Kg	50.0	50.0		100	80 - 120	P
Magnesium	mg/Kg	100	101		101	80 - 120	P
Manganese	mg/Kg	10.0	9.85		98	80 - 120	P
Nickel	mg/Kg	25.0	25.4		102	80 - 120	P
Potassium	mg/Kg	500	495		99	80 - 120	P
Selenium	mg/Kg	100	107		107	80 - 120	P
Silver	mg/Kg	3.8	3.87		102	80 - 120	P
Sodium	mg/Kg	150	133		89	80 - 120	P
Thallium	mg/Kg	100	108		108	80 - 120	P
Vanadium	mg/Kg	15.0	15.1		101	80 - 120	P
Zinc	mg/Kg	10.0	11.0		110	80 - 120	P