

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115437.D **Time Analyzed:** 16:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.483	5.387	5.587	516.630	500.000	3.3
Aroclor-1016-2	5.504	5.408	5.608	505.140	500.000	1.0
Aroclor-1016-3	5.566	5.470	5.670	548.000	500.000	9.6
Aroclor-1016-4	5.662	5.566	5.766	524.030	500.000	4.8
Aroclor-1016-5	5.953	5.858	6.058	455.710	500.000	-8.9
Aroclor-1260-1	7.067	6.971	7.171	461.650	500.000	-7.7
Aroclor-1260-2	7.320	7.225	7.425	485.430	500.000	-2.9
Aroclor-1260-3	7.678	7.582	7.782	494.900	500.000	-1.0
Aroclor-1260-4	7.900	7.803	8.003	458.210	500.000	-8.4
Aroclor-1260-5	8.207	8.110	8.310	478.190	500.000	-4.4
Decachlorobiphenyl	9.925	9.830	10.030	57.880	50.000	15.8
Tetrachloro-m-xylene	4.343	4.248	4.448	59.040	50.000	18.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115437.D **Time Analyzed:** 16:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.628	4.828	501.090	500.000	0.2
Aroclor-1016-2	4.740	4.647	4.847	500.680	500.000	0.1
Aroclor-1016-3	4.913	4.821	5.021	497.520	500.000	-0.5
Aroclor-1016-4	4.956	4.864	5.064	491.000	500.000	-1.8
Aroclor-1016-5	5.167	5.076	5.276	497.280	500.000	-0.5
Aroclor-1260-1	6.197	6.106	6.306	427.800	500.000	-14.4
Aroclor-1260-2	6.385	6.294	6.494	405.140	500.000	-19.0
Aroclor-1260-3	6.537	6.446	6.646	408.990	500.000	-18.2
Aroclor-1260-4	7.009	6.917	7.117	411.240	500.000	-17.8
Aroclor-1260-5	7.250	7.158	7.358	419.440	500.000	-16.1
Decachlorobiphenyl	8.631	8.541	8.741	54.400	50.000	8.8
Tetrachloro-m-xylene	3.643	3.548	3.748	56.810	50.000	13.6