

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115451.D **Time Analyzed:** 21:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.476	5.387	5.587	541.980	500.000	8.4
Aroclor-1016-2	5.498	5.408	5.608	522.140	500.000	4.4
Aroclor-1016-3	5.560	5.470	5.670	558.920	500.000	11.8
Aroclor-1016-4	5.655	5.566	5.766	524.450	500.000	4.9
Aroclor-1016-5	5.947	5.858	6.058	488.530	500.000	-2.3
Aroclor-1260-1	7.061	6.971	7.171	445.760	500.000	-10.8
Aroclor-1260-2	7.315	7.225	7.425	478.860	500.000	-4.2
Aroclor-1260-3	7.672	7.582	7.782	471.570	500.000	-5.7
Aroclor-1260-4	7.894	7.803	8.003	448.670	500.000	-10.3
Aroclor-1260-5	8.200	8.110	8.310	453.730	500.000	-9.3
Decachlorobiphenyl	9.918	9.830	10.030	55.520	50.000	11.0
Tetrachloro-m-xylene	4.337	4.248	4.448	61.450	50.000	22.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115451.D **Time Analyzed:** 21:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.628	4.828	514.390	500.000	2.9
Aroclor-1016-2	4.737	4.647	4.847	513.960	500.000	2.8
Aroclor-1016-3	4.911	4.821	5.021	504.590	500.000	0.9
Aroclor-1016-4	4.954	4.864	5.064	501.110	500.000	0.2
Aroclor-1016-5	5.166	5.076	5.276	522.460	500.000	4.5
Aroclor-1260-1	6.195	6.106	6.306	437.080	500.000	-12.6
Aroclor-1260-2	6.383	6.294	6.494	407.580	500.000	-18.5
Aroclor-1260-3	6.535	6.446	6.646	420.920	500.000	-15.8
Aroclor-1260-4	7.006	6.917	7.117	411.200	500.000	-17.8
Aroclor-1260-5	7.247	7.158	7.358	404.770	500.000	-19.0
Decachlorobiphenyl	8.627	8.541	8.741	50.690	50.000	1.4
Tetrachloro-m-xylene	3.642	3.548	3.748	58.230	50.000	16.5