

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071532.D **Time Analyzed:** 09:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.497	4.397	4.597	568.890	500.000	13.8
Aroclor-1016-2	4.516	4.416	4.616	588.240	500.000	17.6
Aroclor-1016-3	4.574	4.474	4.674	575.670	500.000	15.1
Aroclor-1016-4	4.664	4.564	4.764	587.080	500.000	17.4
Aroclor-1016-5	4.949	4.849	5.049	583.480	500.000	16.7
Aroclor-1260-1	6.043	5.943	6.143	545.740	500.000	9.1
Aroclor-1260-2	6.300	6.200	6.400	572.600	500.000	14.5
Aroclor-1260-3	6.653	6.552	6.752	568.110	500.000	13.6
Aroclor-1260-4	6.868	6.767	6.967	571.620	500.000	14.3
Aroclor-1260-5	7.175	7.075	7.275	575.150	500.000	15.0
Decachlorobiphenyl	8.538	8.437	8.637	49.610	50.000	-0.8
Tetrachloro-m-xylene	3.410	3.311	3.511	54.390	50.000	8.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071532.D **Time Analyzed:** 09:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.867	3.771	3.971	551.470	500.000	10.3
Aroclor-1016-2	3.883	3.788	3.988	570.220	500.000	14.0
Aroclor-1016-3	4.045	3.950	4.150	548.330	500.000	9.7
Aroclor-1016-4	4.094	3.998	4.198	536.310	500.000	7.3
Aroclor-1016-5	4.291	4.195	4.395	546.750	500.000	9.4
Aroclor-1260-1	5.291	5.197	5.397	540.030	500.000	8.0
Aroclor-1260-2	5.480	5.386	5.586	551.820	500.000	10.4
Aroclor-1260-3	5.624	5.530	5.730	537.710	500.000	7.5
Aroclor-1260-4	6.087	5.993	6.193	543.910	500.000	8.8
Aroclor-1260-5	6.332	6.238	6.438	544.540	500.000	8.9
Decachlorobiphenyl	7.651	7.559	7.759	46.210	50.000	-7.6
Tetrachloro-m-xylene	2.854	2.756	2.956	52.200	50.000	4.4