

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071544.D **Time Analyzed:** 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.502	4.397	4.597	562.870	500.000	12.6
Aroclor-1016-2	4.521	4.416	4.616	579.120	500.000	15.8
Aroclor-1016-3	4.579	4.474	4.674	576.200	500.000	15.2
Aroclor-1016-4	4.669	4.564	4.764	580.240	500.000	16.0
Aroclor-1016-5	4.954	4.849	5.049	567.640	500.000	13.5
Aroclor-1260-1	6.048	5.943	6.143	522.490	500.000	4.5
Aroclor-1260-2	6.305	6.200	6.400	560.550	500.000	12.1
Aroclor-1260-3	6.657	6.552	6.752	551.680	500.000	10.3
Aroclor-1260-4	6.873	6.767	6.967	556.820	500.000	11.4
Aroclor-1260-5	7.181	7.075	7.275	561.400	500.000	12.3
Decachlorobiphenyl	8.545	8.437	8.637	48.210	50.000	-3.6
Tetrachloro-m-xylene	3.415	3.311	3.511	52.440	50.000	4.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: REMI01
Lab Code: ACE
SDG NO.: Q3719
GC Column: ZB-MR2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 11/18/2025 11/18/2025

Client Sample No.: CCAL04
Date Analyzed: 11/26/2025
Lab Sample No.: AR1660CCC500
Data File : PQ071544.D
Time Analyzed: 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.866	3.771	3.971	537.160	500.000	7.4
Aroclor-1016-2	3.883	3.788	3.988	561.240	500.000	12.2
Aroclor-1016-3	4.045	3.950	4.150	537.110	500.000	7.4
Aroclor-1016-4	4.094	3.998	4.198	525.180	500.000	5.0
Aroclor-1016-5	4.291	4.195	4.395	535.090	500.000	7.0
Aroclor-1260-1	5.292	5.197	5.397	535.670	500.000	7.1
Aroclor-1260-2	5.481	5.386	5.586	544.500	500.000	8.9
Aroclor-1260-3	5.624	5.530	5.730	536.270	500.000	7.3
Aroclor-1260-4	6.088	5.993	6.193	533.340	500.000	6.7
Aroclor-1260-5	6.333	6.238	6.438	534.980	500.000	7.0
Decachlorobiphenyl	7.653	7.559	7.759	45.360	50.000	-9.3
Tetrachloro-m-xylene	2.853	2.756	2.956	50.570	50.000	1.1