

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071550.D **Time Analyzed:** 18:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.496	4.397	4.597	570.390	500.000	14.1
Aroclor-1016-2	4.515	4.416	4.616	572.560	500.000	14.5
Aroclor-1016-3	4.573	4.474	4.674	562.410	500.000	12.5
Aroclor-1016-4	4.663	4.564	4.764	567.660	500.000	13.5
Aroclor-1016-5	4.948	4.849	5.049	563.070	500.000	12.6
Aroclor-1260-1	6.043	5.943	6.143	506.000	500.000	1.2
Aroclor-1260-2	6.300	6.200	6.400	531.450	500.000	6.3
Aroclor-1260-3	6.652	6.552	6.752	512.530	500.000	2.5
Aroclor-1260-4	6.867	6.767	6.967	520.920	500.000	4.2
Aroclor-1260-5	7.175	7.075	7.275	523.030	500.000	4.6
Decachlorobiphenyl	8.537	8.437	8.637	51.820	50.000	3.6
Tetrachloro-m-xylene	3.410	3.311	3.511	59.940	50.000	19.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/18/2025 11/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071550.D **Time Analyzed:** 18:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.867	3.771	3.971	558.180	500.000	11.6
Aroclor-1016-2	3.883	3.788	3.988	578.690	500.000	15.7
Aroclor-1016-3	4.045	3.950	4.150	556.980	500.000	11.4
Aroclor-1016-4	4.094	3.998	4.198	543.740	500.000	8.7
Aroclor-1016-5	4.291	4.195	4.395	554.260	500.000	10.9
Aroclor-1260-1	5.291	5.197	5.397	528.680	500.000	5.7
Aroclor-1260-2	5.481	5.386	5.586	531.730	500.000	6.3
Aroclor-1260-3	5.623	5.530	5.730	521.230	500.000	4.2
Aroclor-1260-4	6.087	5.993	6.193	517.000	500.000	3.4
Aroclor-1260-5	6.331	6.238	6.438	515.940	500.000	3.2
Decachlorobiphenyl	7.652	7.559	7.759	45.640	50.000	-8.7
Tetrachloro-m-xylene	2.854	2.756	2.956	58.480	50.000	17.0