

Surrogate Summary

SDG No.: Q3719

Client: Remington & Vernick

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114870.D	PIBLK-PO114870.D	Tetrachloro-m-xyl	1	20	15.5	77		60	140
		Decachlorobiphen	1	20	16.8	84		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
		Decachlorobiphen	2	20	17.3	86		60	140
I.BLK-PO115441.D	PIBLK-PO115441.D	Tetrachloro-m-xyl	1	20	25.7	129		60	140
		Decachlorobiphen	1	20	25.3	126		60	140
		Tetrachloro-m-xyl	2	20	25.0	125		60	140
		Decachlorobiphen	2	20	24.1	121		60	140
PB170739BL	PB170739BL	Tetrachloro-m-xyl	1	20	27.9	139		21	165
		Decachlorobiphen	1	20	28.3	141		10	170
		Tetrachloro-m-xyl	2	20	27.1	136		21	165
		Decachlorobiphen	2	20	27.4	137		10	170
PB170739BS	PB170739BS	Tetrachloro-m-xyl	1	20	26.5	133		21	165
		Decachlorobiphen	1	20	26.9	134		10	170
		Tetrachloro-m-xyl	2	20	24.4	122		21	165
		Decachlorobiphen	2	20	25.8	129		10	170
Q3714-01MS	OR-03-11-24-2025MS	Tetrachloro-m-xyl	1	20	17.6	88		21	165
		Decachlorobiphen	1	20	17.6	88		10	170
		Tetrachloro-m-xyl	2	20	20.6	103		21	165
		Decachlorobiphen	2	20	16.6	83		10	170
Q3714-01MSD	OR-03-11-24-2025MSD	Tetrachloro-m-xyl	1	20	19.0	95		21	165
		Decachlorobiphen	1	20	21.1	105		10	170
		Tetrachloro-m-xyl	2	20	23.3	117		21	165
		Decachlorobiphen	2	20	17.4	87		10	170
I.BLK-PO115455.D	PIBLK-PO115455.D	Tetrachloro-m-xyl	1	20	25.8	129		60	140
		Decachlorobiphen	1	20	25.2	126		60	140
		Tetrachloro-m-xyl	2	20	25.1	125		60	140
		Decachlorobiphen	2	20	23.5	117		60	140
I.BLK-PQ071272.D	PIBLK-PQ071272.D	Tetrachloro-m-xyl	1	20	21.9	110		60	140
		Decachlorobiphen	1	20	22.2	111		60	140
		Tetrachloro-m-xyl	2	20	22.2	111		60	140
		Decachlorobiphen	2	20	21.7	109		60	140
I.BLK-PQ071533.D	PIBLK-PQ071533.D	Tetrachloro-m-xyl	1	20	25.8	129		60	140
		Decachlorobiphen	1	20	24.5	123		60	140
		Tetrachloro-m-xyl	2	20	25.1	125		60	140
		Decachlorobiphen	2	20	22.4	112		60	140
Q3719-01	SB-1125-10A	Tetrachloro-m-xyl	1	20	20.7	104		21	165
		Decachlorobiphen	1	20	14.0	70		10	170
		Tetrachloro-m-xyl	2	20	19.5	97		21	165
		Decachlorobiphen	2	20	12.3	62		10	170
Q3719-02	SB-1125-10B	Tetrachloro-m-xyl	1	20	24.5	123		21	165

Surrogate Summary

SDG No.: Q3719

Client: Remington & Vernick

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3719-02	SB-1125-10B	Decachlorobiphen	1	20	18.8	94		10	170
		Tetrachloro-m-xyl	2	20	23.4	117		21	165
Q3719-03	SB-1125-11A	Decachlorobiphen	2	20	16.7	84		10	170
		Tetrachloro-m-xyl	1	20	25.4	127		21	165
		Decachlorobiphen	1	20	20.9	104		10	170
		Tetrachloro-m-xyl	2	20	24.5	123		21	165
		Decachlorobiphen	2	20	18.3	92		10	170
Q3719-04	SB-1125-11B	Tetrachloro-m-xyl	1	20	25.8	129		21	165
		Decachlorobiphen	1	20	23.9	119		10	170
		Tetrachloro-m-xyl	2	20	24.9	125		21	165
		Decachlorobiphen	2	20	21.0	105		10	170
Q3719-05	SB-1125-12A	Tetrachloro-m-xyl	1	20	20.5	102		21	165
		Decachlorobiphen	1	20	12.6	63		10	170
		Tetrachloro-m-xyl	2	20	19.8	99		21	165
		Decachlorobiphen	2	20	11.3	57		10	170
Q3719-06	SB-1125-12B	Tetrachloro-m-xyl	1	20	24.5	123		21	165
		Decachlorobiphen	1	20	21.6	108		10	170
		Tetrachloro-m-xyl	2	20	23.4	117		21	165
		Decachlorobiphen	2	20	17.2	86		10	170
Q3719-07	SB-1125-14A	Tetrachloro-m-xyl	1	20	24.0	120		21	165
		Decachlorobiphen	1	20	20.3	101		10	170
		Tetrachloro-m-xyl	2	20	22.8	114		21	165
		Decachlorobiphen	2	20	18.4	92		10	170
Q3719-08	SB-1125-14B	Tetrachloro-m-xyl	1	20	26.9	135		21	165
		Decachlorobiphen	1	20	25.1	126		10	170
		Tetrachloro-m-xyl	2	20	26.3	131		21	165
		Decachlorobiphen	2	20	22.3	112		10	170
Q3719-09	SB-1125-15A	Tetrachloro-m-xyl	1	20	25.9	130		21	165
		Decachlorobiphen	1	20	18.5	93		10	170
		Tetrachloro-m-xyl	2	20	25.4	127		21	165
		Decachlorobiphen	2	20	17.4	87		10	170
Q3719-10	SB-1125-15B	Tetrachloro-m-xyl	1	20	25.6	128		21	165
		Decachlorobiphen	1	20	20.8	104		10	170
		Tetrachloro-m-xyl	2	20	25.1	125		21	165
		Decachlorobiphen	2	20	18.7	94		10	170
I.BLK-PQ071545.D	PIBLK-PQ071545.D	Tetrachloro-m-xyl	1	20	27.8	139		60	140
		Decachlorobiphen	1	20	26.5	133		60	140
		Tetrachloro-m-xyl	2	20	27.4	137		60	140
		Decachlorobiphen	2	20	24.6	123		60	140
Q3719-11	SB-1125-16A	Tetrachloro-m-xyl	1	20	22.4	112		21	165
		Decachlorobiphen	1	20	15.7	79		10	170

Surrogate Summary

SDG No.: Q3719

Client: Remington & Vernick

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3719-11	SB-1125-16A	Tetrachloro-m-xyl	2	20	21.9	109		21	165
		Decachlorobiphen	2	20	14.2	71		10	170
Q3719-12	SB-1125-16B	Tetrachloro-m-xyl	1	20	27.2	136		21	165
		Decachlorobiphen	1	20	24.6	123		10	170
		Tetrachloro-m-xyl	2	20	26.8	134		21	165
		Decachlorobiphen	2	20	20.8	104		10	170
Q3719-13	SB-1125-17A	Tetrachloro-m-xyl	1	20	20.1	101		21	165
		Decachlorobiphen	1	20	16.1	80		10	170
		Tetrachloro-m-xyl	2	20	22.1	111		21	165
		Decachlorobiphen	2	20	11.1	56		10	170
Q3719-14	SB-1125-17B	Tetrachloro-m-xyl	1	20	22.2	111		21	165
		Decachlorobiphen	1	20	17.1	86		10	170
		Tetrachloro-m-xyl	2	20	23.1	116		21	165
		Decachlorobiphen	2	20	11.2	56		10	170
I.BLK-PQ071551.D	PIBLK-PQ071551.D	Tetrachloro-m-xyl	1	20	26.8	134		60	140
		Decachlorobiphen	1	20	24.1	120		60	140
		Tetrachloro-m-xyl	2	20	26.9	134		60	140
		Decachlorobiphen	2	20	21.6	108		60	140