



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3719	Analytical Method:	8082A
Client:	Remington & Vernick	Datafile :	PO115443.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170739BS (Column 1)	AR1016	166.5	155	ug/kg	93				71	120	
	AR1260	166.5	142	ug/kg	85				65	130	
PB170739BS (Column 2)	AR1016	166.5	145	ug/kg	87				71	120	
	AR1260	166.5	127	ug/kg	76				65	130	