

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/26/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091369.D **Time Analyzed:** 09:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.697	6.599	6.799	48.290	50.000	-3.4
4,4'-DDE	6.187	6.089	6.289	49.080	50.000	-1.8
4,4'-DDT	7.012	6.914	7.114	51.690	50.000	3.4
Aldrin	5.260	5.163	5.363	49.570	50.000	-0.9
alpha-BHC	3.991	3.893	4.093	50.460	50.000	0.9
alpha-Chlordane	6.017	5.919	6.119	48.230	50.000	-3.5
beta-BHC	4.509	4.411	4.611	48.130	50.000	-3.7
Decachlorobiphenyl	9.061	8.963	9.163	47.440	50.000	-5.1
delta-BHC	4.757	4.660	4.860	50.160	50.000	0.3
Dieldrin	6.337	6.239	6.439	49.640	50.000	-0.7
Endosulfan I	6.065	5.966	6.166	49.050	50.000	-1.9
Endosulfan II	6.778	6.679	6.879	48.350	50.000	-3.3
Endosulfan sulfate	7.141	7.043	7.243	48.150	50.000	-3.7
Endrin	6.565	6.466	6.666	51.600	50.000	3.2
Endrin aldehyde	6.907	6.808	7.008	47.900	50.000	-4.2
Endrin ketone	7.622	7.524	7.724	48.200	50.000	-3.6
gamma-BHC (Lindane)	4.322	4.224	4.424	50.210	50.000	0.4
gamma-Chlordane	5.937	5.838	6.038	49.440	50.000	-1.1
Heptachlor	4.919	4.821	5.021	49.680	50.000	-0.6
Heptachlor epoxide	5.681	5.582	5.782	49.380	50.000	-1.2
Methoxychlor	7.486	7.387	7.587	49.310	50.000	-1.4
Tetrachloro-m-xylene	3.542	3.444	3.644	48.590	50.000	-2.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/26/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091369.D **Time Analyzed:** 09:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	52.770	50.000	5.5
4,4'-DDE	5.361	5.263	5.463	55.270	50.000	10.5
4,4'-DDT	6.169	6.071	6.271	58.960	50.000	17.9
Aldrin	4.356	4.258	4.458	54.140	50.000	8.3
alpha-BHC	3.383	3.286	3.486	53.120	50.000	6.2
alpha-Chlordane	5.177	5.079	5.279	54.620	50.000	9.2
beta-BHC	4.015	3.918	4.118	52.200	50.000	4.4
Decachlorobiphenyl	8.056	7.958	8.158	55.520	50.000	11.0
delta-BHC	4.252	4.153	4.353	53.270	50.000	6.5
Dieldrin	5.499	5.401	5.601	54.500	50.000	9.0
Endosulfan I	5.233	5.135	5.335	52.210	50.000	4.4
Endosulfan II	6.067	5.969	6.169	54.830	50.000	9.7
Endosulfan sulfate	6.469	6.371	6.571	54.850	50.000	9.7
Endrin	5.776	5.678	5.878	57.320	50.000	14.6
Endrin aldehyde	6.245	6.148	6.348	53.790	50.000	7.6
Endrin ketone	6.978	6.880	7.080	54.950	50.000	9.9
gamma-BHC (Lindane)	3.719	3.621	3.821	53.240	50.000	6.5
gamma-Chlordane	5.112	5.014	5.214	54.950	50.000	9.9
Heptachlor	4.071	3.973	4.173	52.970	50.000	5.9
Heptachlor epoxide	4.859	4.762	4.962	53.510	50.000	7.0
Methoxychlor	6.739	6.641	6.841	56.350	50.000	12.7
Tetrachloro-m-xylene	2.872	2.774	2.974	52.460	50.000	4.9