

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/26/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091378.D **Time Analyzed:** 15:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.599	6.799	47.380	50.000	-5.2
4,4'-DDE	6.188	6.089	6.289	49.230	50.000	-1.5
4,4'-DDT	7.014	6.914	7.114	52.590	50.000	5.2
Aldrin	5.261	5.163	5.363	52.430	50.000	4.9
alpha-BHC	3.991	3.893	4.093	55.700	50.000	11.4
alpha-Chlordane	6.018	5.919	6.119	47.470	50.000	-5.1
beta-BHC	4.510	4.411	4.611	52.220	50.000	4.4
Decachlorobiphenyl	9.063	8.963	9.163	46.610	50.000	-6.8
delta-BHC	4.758	4.660	4.860	54.280	50.000	8.6
Dieldrin	6.338	6.239	6.439	49.660	50.000	-0.7
Endosulfan I	6.065	5.966	6.166	49.220	50.000	-1.6
Endosulfan II	6.779	6.679	6.879	47.950	50.000	-4.1
Endosulfan sulfate	7.143	7.043	7.243	46.830	50.000	-6.3
Endrin	6.565	6.466	6.666	51.510	50.000	3.0
Endrin aldehyde	6.908	6.808	7.008	47.400	50.000	-5.2
Endrin ketone	7.624	7.524	7.724	47.350	50.000	-5.3
gamma-BHC (Lindane)	4.322	4.224	4.424	55.320	50.000	10.6
gamma-Chlordane	5.937	5.838	6.038	50.060	50.000	0.1
Heptachlor	4.920	4.821	5.021	53.080	50.000	6.2
Heptachlor epoxide	5.682	5.582	5.782	51.380	50.000	2.8
Methoxychlor	7.487	7.387	7.587	48.080	50.000	-3.8
Tetrachloro-m-xylene	3.542	3.444	3.644	53.120	50.000	6.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/26/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091378.D **Time Analyzed:** 15:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.917	5.817	6.017	51.480	50.000	3.0
4,4'-DDE	5.362	5.263	5.463	56.960	50.000	13.9
4,4'-DDT	6.171	6.071	6.271	58.000	50.000	16.0
Aldrin	4.357	4.258	4.458	58.760	50.000	17.5
alpha-BHC	3.383	3.286	3.486	58.320	50.000	16.6
alpha-Chlordane	5.177	5.079	5.279	56.660	50.000	13.3
beta-BHC	4.016	3.918	4.118	56.850	50.000	13.7
Decachlorobiphenyl	8.058	7.958	8.158	43.400	50.000	-13.2
delta-BHC	4.252	4.153	4.353	58.300	50.000	16.6
Dieldrin	5.500	5.401	5.601	55.800	50.000	11.6
Endosulfan I	5.234	5.135	5.335	53.540	50.000	7.1
Endosulfan II	6.069	5.969	6.169	52.830	50.000	5.7
Endosulfan sulfate	6.470	6.371	6.571	50.920	50.000	1.8
Endrin	5.777	5.678	5.878	58.740	50.000	17.5
Endrin aldehyde	6.247	6.148	6.348	51.280	50.000	2.6
Endrin ketone	6.979	6.880	7.080	46.490	50.000	-7.0
gamma-BHC (Lindane)	3.719	3.621	3.821	58.430	50.000	16.9
gamma-Chlordane	5.113	5.014	5.214	57.200	50.000	14.4
Heptachlor	4.071	3.973	4.173	58.780	50.000	17.6
Heptachlor epoxide	4.860	4.762	4.962	56.350	50.000	12.7
Methoxychlor	6.741	6.641	6.841	52.900	50.000	5.8
Tetrachloro-m-xylene	2.872	2.774	2.974	57.140	50.000	14.3