

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/26/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091390.D **Time Analyzed:** 18:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.599	6.799	44.410	50.000	-11.2
4,4'-DDE	6.187	6.089	6.289	47.590	50.000	-4.8
4,4'-DDT	7.013	6.914	7.114	48.430	50.000	-3.1
Aldrin	5.261	5.163	5.363	48.460	50.000	-3.1
alpha-BHC	3.992	3.893	4.093	50.100	50.000	0.2
alpha-Chlordane	6.018	5.919	6.119	44.760	50.000	-10.5
beta-BHC	4.510	4.411	4.611	47.530	50.000	-4.9
Decachlorobiphenyl	9.063	8.963	9.163	42.780	50.000	-14.4
delta-BHC	4.758	4.660	4.860	49.320	50.000	-1.4
Dieldrin	6.338	6.239	6.439	47.220	50.000	-5.6
Endosulfan I	6.065	5.966	6.166	46.880	50.000	-6.2
Endosulfan II	6.779	6.679	6.879	44.610	50.000	-10.8
Endosulfan sulfate	7.142	7.043	7.243	44.310	50.000	-11.4
Endrin	6.565	6.466	6.666	48.860	50.000	-2.3
Endrin aldehyde	6.907	6.808	7.008	43.660	50.000	-12.7
Endrin ketone	7.623	7.524	7.724	44.190	50.000	-11.6
gamma-BHC (Lindane)	4.322	4.224	4.424	49.820	50.000	-0.4
gamma-Chlordane	5.937	5.838	6.038	47.600	50.000	-4.8
Heptachlor	4.920	4.821	5.021	48.970	50.000	-2.1
Heptachlor epoxide	5.681	5.582	5.782	47.230	50.000	-5.5
Methoxychlor	7.486	7.387	7.587	44.480	50.000	-11.0
Tetrachloro-m-xylene	3.542	3.444	3.644	48.240	50.000	-3.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/26/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091390.D **Time Analyzed:** 18:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.916	5.817	6.017	49.680	50.000	-0.6
4,4'-DDE	5.362	5.263	5.463	52.210	50.000	4.4
4,4'-DDT	6.170	6.071	6.271	54.220	50.000	8.4
Aldrin	4.357	4.258	4.458	52.900	50.000	5.8
alpha-BHC	3.384	3.286	3.486	52.470	50.000	4.9
alpha-Chlordane	5.177	5.079	5.279	52.910	50.000	5.8
beta-BHC	4.016	3.918	4.118	51.910	50.000	3.8
Decachlorobiphenyl	8.058	7.958	8.158	43.600	50.000	-12.8
delta-BHC	4.252	4.153	4.353	52.240	50.000	4.5
Dieldrin	5.500	5.401	5.601	52.410	50.000	4.8
Endosulfan I	5.235	5.135	5.335	52.760	50.000	5.5
Endosulfan II	6.069	5.969	6.169	50.460	50.000	0.9
Endosulfan sulfate	6.470	6.371	6.571	49.780	50.000	-0.4
Endrin	5.777	5.678	5.878	54.640	50.000	9.3
Endrin aldehyde	6.247	6.148	6.348	48.790	50.000	-2.4
Endrin ketone	6.979	6.880	7.080	47.750	50.000	-4.5
gamma-BHC (Lindane)	3.720	3.621	3.821	52.660	50.000	5.3
gamma-Chlordane	5.113	5.014	5.214	53.310	50.000	6.6
Heptachlor	4.072	3.973	4.173	52.840	50.000	5.7
Heptachlor epoxide	4.860	4.762	4.962	52.400	50.000	4.8
Methoxychlor	6.741	6.641	6.841	51.550	50.000	3.1
Tetrachloro-m-xylene	2.873	2.774	2.974	51.800	50.000	3.6