

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/26/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091399.D **Time Analyzed:** 20:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.697	6.599	6.799	53.080	50.000	6.2
4,4'-DDE	6.187	6.089	6.289	53.920	50.000	7.8
4,4'-DDT	7.013	6.914	7.114	54.180	50.000	8.4
Aldrin	5.261	5.163	5.363	55.470	50.000	10.9
alpha-BHC	3.991	3.893	4.093	57.420	50.000	14.8
alpha-Chlordane	6.017	5.919	6.119	51.270	50.000	2.5
beta-BHC	4.509	4.411	4.611	54.230	50.000	8.5
Decachlorobiphenyl	9.063	8.963	9.163	47.230	50.000	-5.5
delta-BHC	4.758	4.660	4.860	57.110	50.000	14.2
Dieldrin	6.338	6.239	6.439	54.270	50.000	8.5
Endosulfan I	6.065	5.966	6.166	53.650	50.000	7.3
Endosulfan II	6.778	6.679	6.879	52.140	50.000	4.3
Endosulfan sulfate	7.142	7.043	7.243	50.900	50.000	1.8
Endrin	6.565	6.466	6.666	56.900	50.000	13.8
Endrin aldehyde	6.907	6.808	7.008	50.520	50.000	1.0
Endrin ketone	7.623	7.524	7.724	49.680	50.000	-0.6
gamma-BHC (Lindane)	4.322	4.224	4.424	57.110	50.000	14.2
gamma-Chlordane	5.936	5.838	6.038	54.730	50.000	9.5
Heptachlor	4.919	4.821	5.021	55.480	50.000	11.0
Heptachlor epoxide	5.681	5.582	5.782	53.560	50.000	7.1
Methoxychlor	7.486	7.387	7.587	51.100	50.000	2.2
Tetrachloro-m-xylene	3.542	3.444	3.644	54.420	50.000	8.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3719
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/14/2025 11/14/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/26/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091399.D **Time Analyzed:** 20:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.917	5.817	6.017	55.450	50.000	10.9
4,4'-DDE	5.362	5.263	5.463	58.980	50.000	18.0
4,4'-DDT	6.170	6.071	6.271	59.870	50.000	19.7
Aldrin	4.357	4.258	4.458	59.280	50.000	18.6
alpha-BHC	3.384	3.286	3.486	58.590	50.000	17.2
alpha-Chlordane	5.177	5.079	5.279	58.300	50.000	16.6
beta-BHC	4.016	3.918	4.118	57.120	50.000	14.2
Decachlorobiphenyl	8.058	7.958	8.158	47.880	50.000	-4.2
delta-BHC	4.252	4.153	4.353	58.580	50.000	17.2
Dieldrin	5.500	5.401	5.601	57.540	50.000	15.1
Endosulfan I	5.234	5.135	5.335	54.790	50.000	9.6
Endosulfan II	6.068	5.969	6.169	56.790	50.000	13.6
Endosulfan sulfate	6.470	6.371	6.571	54.910	50.000	9.8
Endrin	5.776	5.678	5.878	61.990	50.000	24.0
Endrin aldehyde	6.246	6.148	6.348	54.450	50.000	8.9
Endrin ketone	6.979	6.880	7.080	51.370	50.000	2.7
gamma-BHC (Lindane)	3.720	3.621	3.821	58.550	50.000	17.1
gamma-Chlordane	5.113	5.014	5.214	58.960	50.000	17.9
Heptachlor	4.071	3.973	4.173	58.160	50.000	16.3
Heptachlor epoxide	4.860	4.762	4.962	57.850	50.000	15.7
Methoxychlor	6.740	6.641	6.841	55.480	50.000	11.0
Tetrachloro-m-xylene	2.872	2.774	2.974	57.380	50.000	14.8