

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112625\
 Data File : PD091369.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2025 09:52
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 11:08:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 15 04:46:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.872	167.6E6	1921.7E6	48.586	52.461
28)	SA Decachlor...	9.061	8.056	236.4E6	2197.3E6	47.443	55.523
Target Compounds							
2)	A alpha-BHC	3.991	3.383	403.5E6	3357.3E6	50.462	53.119
3)	MA gamma-BHC...	4.322	3.719	375.2E6	3078.4E6	50.215	53.239
4)	MA Heptachlor	4.919	4.071	355.8E6	2924.4E6	49.681	52.968
5)	MB Aldrin	5.260	4.356	359.7E6	3056.6E6	49.565	54.139
6)	B beta-BHC	4.509	4.015	139.8E6	1254.1E6	48.128	52.195
7)	B delta-BHC	4.757	4.252	372.8E6	3115.1E6	50.157	53.271
8)	B Heptachlo...	5.681	4.859	320.8E6	2726.9E6	49.380	53.512
9)	A Endosulfan I	6.065	5.233	301.3E6	2498.2E6	49.052	52.208
10)	B gamma-Chl...	5.937	5.112	321.4E6	3001.8E6	49.438	54.951
11)	B alpha-Chl...	6.017	5.177	332.0E6	2858.3E6	48.231	54.619
12)	B 4,4'-DDE	6.187	5.361	284.7E6	2953.8E6	49.084	55.272
13)	MA Dieldrin	6.337	5.499	320.9E6	2960.1E6	49.638	54.503
14)	MA Endrin	6.565	5.776	256.1E6	2590.5E6	51.596	57.322
15)	B Endosulfa...	6.778	6.067	268.8E6	2559.5E6	48.350	54.830
16)	A 4,4'-DDD	6.697	5.916	237.2E6	2578.7E6	48.286	52.768
17)	MA 4,4'-DDT	7.012	6.169	213.8E6	2307.2E6	51.694	58.959
18)	B Endrin al...	6.907	6.245	199.5E6	1898.3E6	47.899	53.794
19)	B Endosulfa...	7.141	6.469	247.7E6	2459.0E6	48.146	54.846
20)	A Methoxychlor	7.486	6.739	107.2E6	1129.8E6	49.308	56.346
21)	B Endrin ke...	7.622	6.978	271.5E6	2800.1E6	48.202	54.952
22)	Mirex	8.104	7.170	170.1E6	1833.3E6	47.814	55.320

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD112625\
Data File : PD091369.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2025 09:52
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 11:08:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD111525.M
Quant Title : GC Extractables
QLast Update : Sat Nov 15 04:46:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

