

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110525

Review By	rahul	Review On	11/5/2025 2:56:50 PM
Supervise By	Jagrut	Supervise On	11/6/2025 10:21:33 AM
SubDirectory	BF110525	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144168.D	05 Nov 2025 12:21		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF144169.D	05 Nov 2025 12:50		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF144170.D	05 Nov 2025 13:20		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF144171.D	05 Nov 2025 13:50		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF144172.D	05 Nov 2025 14:20		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF144173.D	05 Nov 2025 14:50	Compound #41 kept on QR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF144174.D	05 Nov 2025 15:49		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF144175.D	05 Nov 2025 16:19		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF144176.D	05 Nov 2025 16:49		RC/JU	Ok
10	SSTDICV040	ICVBF110525	BF144177.D	05 Nov 2025 17:21		RC/JU	Ok,M
11	PB170372BL	PB170372BL	BF144178.D	05 Nov 2025 18:10		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112625

Review By	Jagrut	Review On	12/2/2025 4:03:33 PM
Supervise By	mohammad	Supervise On	12/3/2025 1:12:43 AM
SubDirectory	BF112625	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13184,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144357.D	26 Nov 2025 13:00		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144358.D	26 Nov 2025 13:30		RC/JU	Ok
3	PB170735BL	PB170735BL	BF144359.D	26 Nov 2025 14:00		RC/JU	Ok
4	PB170735BS	PB170735BS	BF144360.D	26 Nov 2025 14:30		RC/JU	Ok,M
5	Q3604-04	S-1	BF144361.D	26 Nov 2025 15:11		RC/JU	Ok
6	Q3604-05	S-2	BF144362.D	26 Nov 2025 15:41		RC/JU	Ok
7	Q3719-01	SB-1125-10A	BF144363.D	26 Nov 2025 16:11		RC/JU	Ok,M
8	Q3719-02	SB-1125-10B	BF144364.D	26 Nov 2025 16:42		RC/JU	Ok,M
9	Q3719-03	SB-1125-11A	BF144365.D	26 Nov 2025 17:12		RC/JU	Ok,M
10	Q3719-04	SB-1125-11B	BF144366.D	26 Nov 2025 17:42		RC/JU	Ok,M
11	Q3719-05	SB-1125-12A	BF144367.D	26 Nov 2025 18:12		RC/JU	Ok,M
12	Q3719-06	SB-1125-12B	BF144368.D	26 Nov 2025 18:42		RC/JU	Ok,M
13	Q3719-07	SB-1125-14A	BF144369.D	26 Nov 2025 19:13		RC/JU	Ok,M
14	Q3719-08	SB-1125-14B	BF144370.D	26 Nov 2025 19:43		RC/JU	Ok,M
15	Q3719-09	SB-1125-15A	BF144371.D	26 Nov 2025 20:13		RC/JU	Ok,M
16	Q3719-10	SB-1125-15B	BF144372.D	26 Nov 2025 20:43	Re run	RC/JU	Not Ok
17	Q3719-10MS	SB-1125-15BMS	BF144373.D	26 Nov 2025 21:13	MSD is not ok	RC/JU	Not Ok
18	Q3719-10MSD	SB-1125-15BMSD	BF144374.D	26 Nov 2025 21:43	Many compounds recovery Low	RC/JU	Not Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112625

Review By	Jagrut	Review On	12/2/2025 4:03:33 PM		
Supervise By	mohammad	Supervise On	12/3/2025 1:12:43 AM		
SubDirectory	BF112625	HP Acquire Method	BNA_F	HP Processing Method	BF110525
STD. NAME	STD REF.#				
Tune/Reschk	SP6875				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13184,10ul/1000ul sample				
ICV/I.BLK	SP6872				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3719-12	SB-1125-16B	BF144375.D	26 Nov 2025 22:13		RC/JU	Ok,M
20	Q3714-01	OR-03-11-24-2025	BF144376.D	26 Nov 2025 22:43		RC/JU	Ok,M
21	Q3721-01	EO-2-112525	BF144377.D	26 Nov 2025 23:14		RC/JU	Ok
22	Q3721-01MS	EO-2-112525MS	BF144378.D	26 Nov 2025 23:44		RC/JU	Ok
23	Q3721-01MSD	EO-2-112525MSD	BF144379.D	27 Nov 2025 00:14		RC/JU	Ok,M
24	Q3723-01	CC-1-112525	BF144380.D	27 Nov 2025 00:44		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF120125

Review By	Jagrut	Review On	12/2/2025 4:27:26 PM
Supervise By	mohammad	Supervise On	12/3/2025 1:12:18 AM
SubDirectory	BF120125	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13184,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144381.D	01 Dec 2025 12:12		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144382.D	01 Dec 2025 12:41	CCC fail high for compound #9 & #77	RC/JU	Ok,M
3	PB170749BL	PB170749BL	BF144383.D	01 Dec 2025 13:13		RC/JU	Ok
4	PB170749BS	PB170749BS	BF144384.D	01 Dec 2025 13:42		RC/JU	Ok
5	PB170763BL	PB170763BL	BF144385.D	01 Dec 2025 14:11		RC/JU	Ok
6	PB170763BS	PB170763BS	BF144386.D	01 Dec 2025 14:40		RC/JU	Ok
7	Q3740-01	WC1	BF144387.D	01 Dec 2025 15:15		RC/JU	Ok,M
8	Q3719-11	SB-1125-16A	BF144388.D	01 Dec 2025 15:45		RC/JU	Ok,M
9	Q3725-04	STOCKPILE-SAMPLES	BF144389.D	01 Dec 2025 16:14		RC/JU	Ok,M
10	Q3732-01	HD-01-11-26-2025	BF144390.D	01 Dec 2025 16:44		RC/JU	Ok
11	Q3735-01	BU-3-112625	BF144391.D	01 Dec 2025 17:13		RC/JU	Ok
12	Q3735-01MS	BU-3-112625MS	BF144392.D	01 Dec 2025 17:43		RC/JU	Ok
13	Q3735-01MSD	BU-3-112625MSD	BF144393.D	01 Dec 2025 18:12		RC/JU	Ok
14	Q3719-13	SB-1125-17A	BF144394.D	01 Dec 2025 18:41	Need Straight Run	RC/JU	Not Ok
15	Q3719-14	SB-1125-17B	BF144395.D	01 Dec 2025 19:11		RC/JU	Ok,M
16	Q3733-01	SU-04-11-26-2025	BF144396.D	01 Dec 2025 19:40	Double IS added	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF120225

Review By	Jagrut	Review On	12/3/2025 10:09:52 AM
Supervise By	Sohil	Supervise On	12/3/2025 3:31:56 PM
SubDirectory	BF120225	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13184,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144397.D	02 Dec 2025 10:31		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144398.D	02 Dec 2025 11:00		RC/JU	Ok
3	PB170788BL	PB170788BL	BF144399.D	02 Dec 2025 13:16		RC/JU	Ok
4	PB170788BS	PB170788BS	BF144400.D	02 Dec 2025 13:47		RC/JU	Ok
5	Q3742-01	SB-1125-23	BF144401.D	02 Dec 2025 14:21		RC/JU	Ok,M
6	Q3742-03	SB-1125-8	BF144402.D	02 Dec 2025 14:51	Hit of # 09 (Below CRQL)	RC/JU	Not Ok
7	Q3742-07	SB-1125-18	BF144403.D	02 Dec 2025 15:20		RC/JU	Ok,M
8	Q3742-08	SB-1125-20	BF144404.D	02 Dec 2025 15:50		RC/JU	Ok,M
9	Q3742-09	SB-1125-22	BF144405.D	02 Dec 2025 16:19		RC/JU	Ok,M
10	Q3742-17	SB-1125-25	BF144406.D	02 Dec 2025 16:49		RC/JU	Ok,M
11	Q3742-18	SB-1125-24	BF144407.D	02 Dec 2025 17:19		RC/JU	Ok,M
12	Q3742-16	SB-1125-1	BF144408.D	02 Dec 2025 17:49		RC/JU	Ok
13	Q3746-01	ROW	BF144409.D	02 Dec 2025 18:19		RC/JU	Ok
14	Q3750-01	FENCE WC-1	BF144410.D	02 Dec 2025 18:49	Need 5X Dilution	RC/JU	Dilution
15	Q3750-04	FENCE WC-2	BF144411.D	02 Dec 2025 19:18		RC/JU	Ok,M
16	Q3748-01	AU-05-12-1-2025	BF144412.D	02 Dec 2025 19:48	Internal Standard Fail	RC/JU	Ok
17	Q3719-13	SB-1125-17A	BF144413.D	02 Dec 2025 20:18	Internal Standard Fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP111825

Review By	Jagrut	Review On	11/19/2025 10:53:24 AM
Supervise By	Sohil	Supervise On	11/19/2025 3:40:33 PM
SubDirectory	BP111825	HP Acquire Method	BNA_P
		HP Processing Method	BP111825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6910,SP6911,SP6912,SP6913,SP6914,SP6915,SP6916,SP6917
CCC	SP6913
Internal Standard/PEM	S13183,10ul/1000ul sample
ICV/I.BLK	SP6928
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026142.D	18 Nov 2025 13:45		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP026143.D	18 Nov 2025 14:31		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP026144.D	18 Nov 2025 15:12		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BP026145.D	18 Nov 2025 15:54		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BP026146.D	18 Nov 2025 16:35		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP026147.D	18 Nov 2025 17:57		RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP026148.D	18 Nov 2025 18:38		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BP026149.D	18 Nov 2025 20:00		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP026150.D	18 Nov 2025 21:22		RC/JU	Ok,M
10	SSTDICV040	ICVBP111825	BP026151.D	18 Nov 2025 23:25		RC/JU	Ok,M
11	PB170493TB	PB170493TB	BP026152.D	19 Nov 2025 00:06		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP120125

Review By	Jagrut	Review On	12/2/2025 4:18:26 PM
Supervise By	mohammad	Supervise On	12/3/2025 1:12:23 AM
SubDirectory	BP120125	HP Acquire Method	BNA_P
		HP Processing Method	BP111825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6910,SP6911,SP6912,SP6913,SP6914,SP6915,SP6916,SP6917
CCC	SP6913
Internal Standard/PEM	S13184,10ul/1000ul sample
ICV/I.BLK	SP6928
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026196.D	01 Dec 2025 12:48		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP026197.D	01 Dec 2025 14:09		RC/JU	Ok,M
3	PB170743BL	PB170743BL	BP026198.D	01 Dec 2025 14:50		RC/JU	Ok
4	PB170743BS	PB170743BS	BP026199.D	01 Dec 2025 15:31		RC/JU	Ok,M
5	PB170743BSD	PB170743BSD	BP026200.D	01 Dec 2025 16:12		RC/JU	Ok
6	Q3719-10MS	SB-1125-15BMS	BP026201.D	01 Dec 2025 17:51		RC/JU	Ok
7	Q3719-10MSD	SB-1125-15BMSD	BP026202.D	01 Dec 2025 18:40		RC/JU	Ok
8	Q3719-10	SB-1125-15B	BP026203.D	01 Dec 2025 19:21		RC/JU	Ok
9	Q3716-01	MW5	BP026204.D	01 Dec 2025 20:02		RC/JU	Ok,M
10	Q3716-02	MW3	BP026205.D	01 Dec 2025 20:43		RC/JU	Ok

M : Manual Integration