

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3719

Client ID : SSTDCCC040

Date Analyzed: 12/01/2025

Lab File ID: BF144382.D

Time Analyzed: 12:41

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	58717	6.863	221118	8.15	119295	9.90
UPPER LIMIT	117434	7.363	442236	8.645	238590	10.404
LOWER LIMIT	29358.5	6.363	110559	7.645	59647.5	9.404
EPA SAMPLE NO.						
01 SB-1125-16A	47886	6.86	176365	8.15	87150	9.90
02 SB-1125-17B	64196	6.86	226428	8.15	117202	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.