

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3719

Client ID : SSTDCCC040

Date Analyzed: 12/02/2025

Lab File ID: BF144398.D

Time Analyzed: 11:00

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	57181	6.863	207500	8.15	104779	9.90
UPPER LIMIT	114362	7.363	415000	8.645	209558	10.404
LOWER LIMIT	28590.5	6.363	103750	7.645	52389.5	9.404
EPA SAMPLE NO.						
01 SB-1125-17A	38060	6.86	134068	8.15	71503	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.