

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3719

Client ID : SSTDCCC040

Date Analyzed: 12/01/2025

Lab File ID: BP026197.D

Time Analyzed: 14:09

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	271172	7.66	1093510	10.43	683760	14.29
UPPER LIMIT	542344	8.16	2187020	10.925	1367520	14.789
LOWER LIMIT	135586	7.16	546755	9.925	341880	13.789
EPA SAMPLE NO.						
01 SB-1125-15BMS	233034	7.66	911777	10.41	569978	14.29
02 SB-1125-15B	219787	7.66	815242	10.43	475411	14.28
03 SB-1125-15BMSD	222842	7.64	845732	10.40	491065	14.28

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.