

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP026196.D
Instrument ID: BNA_P

Contract: REMI01
SDG NO.: Q3719
DFTPP Injection Date: 12/01/2025
DFTPP Injection Time: 12:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	31
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	13.2
442	Greater than 50% of mass 198	85.8
443	15.0 - 24.0% of mass 442	17 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP026197.D	12/01/2025	14:09
SB-1125-15BMS	Q3719-10MS	BP026201.D	12/01/2025	17:51
SB-1125-15BMSD	Q3719-10MSD	BP026202.D	12/01/2025	18:40
SB-1125-15B	Q3719-10	BP026203.D	12/01/2025	19:21