284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: REMI01

Lab Code: ACE SDG No.: Q3719

Instrument ID: MSVOA_Y Calibration Date(s): 11/26/2025 11/26/2025

Heated Purge: (Y/N) Y Calibration Time(s): 08:04 10:02

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID: RRF005 = VY023809.D RRF010 = VY023810.D RRF020 = VY023811.D
RRF050 = VY023812.D RRF100 = VY023813.D RRF150 = VY023814.D

COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	$\overline{RRF}$	% RSD
Dichlorodifluoromethane	0.422	0.412	0.414	0.340	0.339	0.334	0.377	11.4
Chloromethane	0.640	0.632	0.607	0.531	0.532	0.515	0.576	9.8
Vinyl Chloride	0.662	0.655	0.640	0.604	0.602	0.580	0.624	5.3
Bromomethane	0.543	0.480	0.439	0.386	0.402	0.390	0.440	14
Chloroethane	0.437	0.420	0.420	0.395	0.394	0.372	0.406	5.8
Trichlorofluoromethane	0.906	0.865	0.874	0.815	0.805	0.793	0.843	5.3
1,1,2-Trichlorotrifluoroethane	0.537	0.509	0.501	0.473	0.477	0.455	0.492	6
1,1-Dichloroethene	0.508	0.487	0.491	0.471	0.476	0.471	0.484	3
Acetone	0.148	0.138	0.137	0.135	0.150	0.133	0.140	5.1
Carbon Disulfide	1.644	1.610	1.618	1.519	1.521	1.461	1.562	4.6
Methyl tert-butyl Ether	1.182	1.158	1.200	1.242	1.355	1.287	1.237	6
Methyl Acetate	0.273	0.264	0.262	0.257	0.299	0.270	0.271	5.5
Methylene Chloride	0.755	0.662	0.582	0.530	0.530	0.500	0.593	16.5
trans-1,2-Dichloroethene	0.553	0.537	0.543	0.526	0.539	0.518	0.536	2.3
1,1-Dichloroethane	1.012	0.982	0.989	0.955	0.968	0.947	0.976	2.4
Cyclohexane	1.060	0.953	0.937	0.896	0.903	0.867	0.936	7.2
2-Butanone	0.169	0.163	0.163	0.169	0.192	0.171	0.171	6.4
Carbon Tetrachloride	0.514	0.490	0.502	0.494	0.507	0.483	0.498	2.3
cis-1,2-Dichloroethene	0.636	0.620	0.623	0.611	0.632	0.613	0.622	1.6
Bromochloromethane	0.437	0.400	0.405	0.399	0.408	0.386	0.406	4.2
Chloroform	1.025	0.991	0.996	0.961	0.969	0.943	0.981	3
1,1,1-Trichloroethane	0.893	0.885	0.882	0.851	0.869	0.844	0.871	2.3
Methylcyclohexane	0.598	0.572	0.591	0.604	0.631	0.603	0.600	3.2
Benzene	1.415	1.388	1.434	1.383	1.410	1.355	1.398	2
1,2-Dichloroethane	0.365	0.354	0.358	0.359	0.372	0.354	0.360	2
Trichloroethene	0.366	0.347	0.371	0.355	0.362	0.350	0.359	2.6
1,2-Dichloropropane	0.324	0.325	0.336	0.329	0.339	0.324	0.330	2
Bromodichloromethane	0.468	0.460	0.468	0.463	0.484	0.460	0.467	1.9
4-Methyl-2-Pentanone	0.183	0.185	0.199	0.219	0.252	0.223	0.210	12.6
Toluene	0.818	0.831	0.883	0.892	0.908	0.866	0.866	4.1

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: REMI01

Lab Code: ACE SDG No.: Q3719

Instrument ID: MSVOA_Y Calibration Date(s): 11/26/2025 11/26/2025

Heated Purge: (Y/N) Y Calibration Time(s): 08:04 10:02

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID: RRF005 = VY023809.D RRF010 = VY023810.D RRF020 = VY023811.D
RRF050 = VY023812.D RRF100 = VY023813.D RRF150 = VY023814.D

COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	$\overline{RRF}$	% RSD
t-1,3-Dichloropropene	0.399	0.394	0.415	0.438	0.473	0.450	0.428	7.2
cis-1,3-Dichloropropene	0.488	0.489	0.509	0.516	0.548	0.526	0.512	4.5
1,1,2-Trichloroethane	0.233	0.230	0.235	0.242	0.257	0.240	0.240	4
2-Hexanone	0.131	0.135	0.147	0.162	0.190	0.167	0.155	14.2
Dibromochloromethane	0.305	0.308	0.321	0.322	0.347	0.326	0.321	4.7
1,2-Dibromoethane	0.218	0.207	0.219	0.220	0.243	0.226	0.222	5.4
Tetrachloroethene	0.525	0.455	0.510	0.470	0.442	0.445	0.474	7.4
Chlorobenzene	1.100	1.091	1.118	1.093	1.109	1.074	1.098	1.4
Ethyl Benzene	1.779	1.789	1.887	1.918	1.967	1.907	1.875	4
m/p-Xylenes	0.676	0.695	0.739	0.743	0.754	0.728	0.722	4.2
o-Xylene	0.633	0.643	0.679	0.690	0.724	0.695	0.677	5
Styrene	0.995	1.047	1.117	1.162	1.210	1.163	1.116	7.2
Bromoform	0.214	0.209	0.209	0.216	0.237	0.221	0.217	4.9
Isopropylbenzene	3.390	3.431	3.571	3.610	3.701	3.590	3.549	3.3
1,1,2,2-Tetrachloroethane	0.580	0.573	0.524	0.576	0.654	0.599	0.584	7.2
1,3-Dichlorobenzene	1.696	1.674	1.670	1.654	1.698	1.645	1.673	1.3
1,4-Dichlorobenzene	1.798	1.678	1.645	1.635	1.662	1.601	1.670	4.1
1,2-Dichlorobenzene	1.489	1.439	1.456	1.460	1.501	1.431	1.463	1.9
1,2-Dibromo-3-Chloropropane	0.104	0.094	0.085	0.095	0.110	0.100	0.098	8.7
1,2,4-Trichlorobenzene	0.855	0.777	0.835	0.877	0.949	0.940	0.872	7.5
1,2,3-Trichlorobenzene	0.711	0.658	0.710	0.740	0.817	0.820	0.743	8.7
1,2-Dichloroethane-d4	0.508	0.478	0.483	0.492	0.486	0.475	0.487	2.5
Dibromofluoromethane	0.299	0.285	0.297	0.303	0.299	0.295	0.296	2.1
Toluene-d8	1.161	1.125	1.183	1.212	1.209	1.170	1.176	2.8
4-Bromofluorobenzene	0.388	0.362	0.381	0.395	0.406	0.390	0.387	3.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.