

Report of Analysis

Client: Remington & Vernick
Project: Saddler Property
Client Sample ID: VW1125SBS01
Lab Sample ID: VW1125SBS01
Analytical Method: 8260D
Sample Wt/Vol: 5 g

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3719
Matrix: SOIL
% Solid: 100
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	22.6		1	1.10	5.00	ug/Kg	11/25/25 10:25	VW112525
74-87-3	Chloromethane	22.3		1	1.10	5.00	ug/Kg	11/25/25 10:25	VW112525
75-01-4	Vinyl Chloride	22.9		1	0.79	5.00	ug/Kg	11/25/25 10:25	VW112525
74-83-9	Bromomethane	21.8		1	1.10	5.00	ug/Kg	11/25/25 10:25	VW112525
75-00-3	Chloroethane	21.9		1	1.30	5.00	ug/Kg	11/25/25 10:25	VW112525
75-69-4	Trichlorofluoromethane	19.0		1	1.20	5.00	ug/Kg	11/25/25 10:25	VW112525
76-13-1	1,1,2-Trichlorotrifluoroethane	23.4		1	1.10	5.00	ug/Kg	11/25/25 10:25	VW112525
75-35-4	1,1-Dichloroethene	22.9		1	1.00	5.00	ug/Kg	11/25/25 10:25	VW112525
67-64-1	Acetone	110		1	4.70	25.0	ug/Kg	11/25/25 10:25	VW112525
75-15-0	Carbon Disulfide	21.2		1	1.10	5.00	ug/Kg	11/25/25 10:25	VW112525
1634-04-4	Methyl tert-butyl Ether	20.6		1	0.73	5.00	ug/Kg	11/25/25 10:25	VW112525
79-20-9	Methyl Acetate	21.8		1	1.50	5.00	ug/Kg	11/25/25 10:25	VW112525
75-09-2	Methylene Chloride	23.9		1	3.50	10.0	ug/Kg	11/25/25 10:25	VW112525
156-60-5	trans-1,2-Dichloroethene	22.1		1	0.86	5.00	ug/Kg	11/25/25 10:25	VW112525
75-34-3	1,1-Dichloroethane	23.4		1	0.80	5.00	ug/Kg	11/25/25 10:25	VW112525
110-82-7	Cyclohexane	22.1		1	0.79	5.00	ug/Kg	11/25/25 10:25	VW112525
78-93-3	2-Butanone	100		1	6.50	25.0	ug/Kg	11/25/25 10:25	VW112525
56-23-5	Carbon Tetrachloride	23.9		1	0.97	5.00	ug/Kg	11/25/25 10:25	VW112525
156-59-2	cis-1,2-Dichloroethene	22.2		1	0.75	5.00	ug/Kg	11/25/25 10:25	VW112525
74-97-5	Bromochloromethane	22.5		1	1.20	5.00	ug/Kg	11/25/25 10:25	VW112525
67-66-3	Chloroform	23.4		1	0.84	5.00	ug/Kg	11/25/25 10:25	VW112525
71-55-6	1,1,1-Trichloroethane	22.9		1	0.93	5.00	ug/Kg	11/25/25 10:25	VW112525
108-87-2	Methylcyclohexane	22.4		1	0.91	5.00	ug/Kg	11/25/25 10:25	VW112525
71-43-2	Benzene	23.6		1	0.79	5.00	ug/Kg	11/25/25 10:25	VW112525
107-06-2	1,2-Dichloroethane	24.1		1	0.79	5.00	ug/Kg	11/25/25 10:25	VW112525
79-01-6	Trichloroethene	22.5		1	0.81	5.00	ug/Kg	11/25/25 10:25	VW112525
78-87-5	1,2-Dichloropropane	24.4		1	0.91	5.00	ug/Kg	11/25/25 10:25	VW112525
75-27-4	Bromodichloromethane	23.6		1	0.78	5.00	ug/Kg	11/25/25 10:25	VW112525
108-10-1	4-Methyl-2-Pentanone	110		1	3.60	25.0	ug/Kg	11/25/25 10:25	VW112525
108-88-3	Toluene	23.9		1	0.78	5.00	ug/Kg	11/25/25 10:25	VW112525
10061-02-6	t-1,3-Dichloropropene	21.7		1	0.65	5.00	ug/Kg	11/25/25 10:25	VW112525
10061-01-5	cis-1,3-Dichloropropene	22.4		1	0.62	5.00	ug/Kg	11/25/25 10:25	VW112525
79-00-5	1,1,2-Trichloroethane	23.1		1	0.92	5.00	ug/Kg	11/25/25 10:25	VW112525
591-78-6	2-Hexanone	110		1	3.70	25.0	ug/Kg	11/25/25 10:25	VW112525
124-48-1	Dibromochloromethane	22.4		1	0.87	5.00	ug/Kg	11/25/25 10:25	VW112525
106-93-4	1,2-Dibromoethane	23.2		1	0.88	5.00	ug/Kg	11/25/25 10:25	VW112525
127-18-4	Tetrachloroethene	22.9		1	1.10	5.00	ug/Kg	11/25/25 10:25	VW112525
108-90-7	Chlorobenzene	22.7		1	0.91	5.00	ug/Kg	11/25/25 10:25	VW112525
100-41-4	Ethyl Benzene	22.1		1	0.67	5.00	ug/Kg	11/25/25 10:25	VW112525
179601-23-1	m/p-Xylenes	46.4		1	1.20	10.0	ug/Kg	11/25/25 10:25	VW112525
95-47-6	o-Xylene	21.5		1	0.82	5.00	ug/Kg	11/25/25 10:25	VW112525
100-42-5	Styrene	22.4		1	0.71	5.00	ug/Kg	11/25/25 10:25	VW112525
75-25-2	Bromoform	21.0		1	0.86	5.00	ug/Kg	11/25/25 10:25	VW112525

Report of Analysis

Client:	Remington & Vernick	Date Collected:	
Project:	Saddler Property	Date Received:	
Client Sample ID:	VW1125SBS01	SDG No.:	Q3719
Lab Sample ID:	VW1125SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 g	Test:	VOC-TCLVOA-10
	Level: LOW		
	Final Vol: 5000 uL		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	20.4		1	0.78	5.00	ug/Kg	11/25/25 10:25	VW112525
79-34-5	1,1,2,2-Tetrachloroethane	21.2		1	1.20	5.00	ug/Kg	11/25/25 10:25	VW112525
541-73-1	1,3-Dichlorobenzene	22.2		1	1.70	5.00	ug/Kg	11/25/25 10:25	VW112525
106-46-7	1,4-Dichlorobenzene	21.3		1	1.60	5.00	ug/Kg	11/25/25 10:25	VW112525
95-50-1	1,2-Dichlorobenzene	20.8		1	1.50	5.00	ug/Kg	11/25/25 10:25	VW112525
96-12-8	1,2-Dibromo-3-Chloropropane	18.4		1	1.80	5.00	ug/Kg	11/25/25 10:25	VW112525
120-82-1	1,2,4-Trichlorobenzene	19.0		1	3.00	5.00	ug/Kg	11/25/25 10:25	VW112525
87-61-6	1,2,3-Trichlorobenzene	19.6		1	3.20	5.00	ug/Kg	11/25/25 10:25	VW112525
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	52.9			63 - 155	106%	SPK: 50		
1868-53-7	Dibromofluoromethane	55.2			70 - 134	110%	SPK: 50		
2037-26-5	Toluene-d8	54.1			74 - 123	108%	SPK: 50		
460-00-4	4-Bromofluorobenzene	54.4			52 - 138	109%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	348000							
540-36-3	1,4-Difluorobenzene	621000							
3114-55-4	Chlorobenzene-d5	580000							
3855-82-1	1,4-Dichlorobenzene-d4	287000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products