

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112625\
 Data File : VW032547.D
 Acq On : 26 Nov 2025 16:27
 Operator : SY/MD
 Sample : Q3719-08
 Misc : 3.72g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-1125-14B

Quant Time: Nov 27 00:29:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	191327	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	462395	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	513156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	246111	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	171946	69.708	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	139.420%
35) Dibromofluoromethane	7.904	113	152930	55.240	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	110.480%
50) Toluene-d8	10.325	98	500524	47.127	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	94.260%
62) 4-Bromofluorobenzene	12.617	95	222080	55.352	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	110.700%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112625\
Data File : VW032547.D
Acq On : 26 Nov 2025 16:27
Operator : SY/MD
Sample : Q3719-08
Misc : 3.72g/5mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-1125-14B

Quant Time: Nov 27 00:29:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.965 min Scan# 997
Delta R.T. 0.000 min
Lab File: VW032547.D
Acq: 26 Nov 2025 16:27




Instrument :
MSVOA_W
ClientSampleId :
SB-1125-14B


Tgt Ion:168 Resp: 191327
Ion Ratio Lower Upper
168 100
99 61.1 45.4 68.2



#33
1,2-Dichloroethane-d4
Concen: 69.708 ug/l
RT: 8.319 min Scan# 1055
Delta R.T. 0.000 min
Lab File: VW032547.D
Acq: 26 Nov 2025 16:27

Tgt Ion: 65 Resp: 171946
Ion Ratio Lower Upper
65 100
67 51.0 0.0 106.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.849 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VW032547.D
Acq: 26 Nov 2025 16:27




Instrument :
MSVOA_W
ClientSampleId :
SB-1125-14B


Tgt Ion:114 Resp: 462395
Ion Ratio Lower Upper
114 100
63 22.3 0.0 41.8
88 16.1 0.0 32.6



#35
Dibromofluoromethane
Concen: 55.240 ug/l
RT: 7.904 min Scan# 987
Delta R.T. 0.000 min
Lab File: VW032547.D
Acq: 26 Nov 2025 16:27

Tgt Ion:113 Resp: 152930
Ion Ratio Lower Upper
113 100
111 104.8 83.1 124.7
192 16.1 13.4 20.2









#50
Toluene-d8
Concen: 47.127 ug/l
RT: 10.325 min Scan# 1
Delta R.T. 0.000 min
Lab File: VW032547.D
Acq: 26 Nov 2025 16:27




Instrument :
MSVOA_W
ClientSampleId :
SB-1125-14B


Tgt Ion: 98 Resp: 500524
Ion Ratio Lower Upper
98 100
100 63.3 53.3 79.9



#62
4-Bromofluorobenzene
Concen: 55.352 ug/l
RT: 12.617 min Scan# 1760
Delta R.T. 0.000 min
Lab File: VW032547.D
Acq: 26 Nov 2025 16:27

Tgt Ion: 95 Resp: 222080
Ion Ratio Lower Upper
95 100
174 67.5 0.0 135.4
176 63.4 0.0 131.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.635 min Scan# 11

Delta R.T. 0.000 min

Lab File: VW032547.D

Acq: 26 Nov 2025 16:27

Instrument :

MSVOA_W

ClientSampleId :

SB-1125-14B





Tgt Ion:117 Resp: 513156

Ion Ratio Lower Upper

117 100

82 59.4 43.0 64.4

119 32.2 25.0 37.6



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.556 min Scan# 1914

Delta R.T. 0.000 min

Lab File: VW032547.D

Acq: 26 Nov 2025 16:27

Tgt Ion:152 Resp: 246111

Ion Ratio Lower Upper

152 100

115 63.0 32.8 98.3

150 150.8 0.0 356.6

