

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3719</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Remington & Vernick</u>	Datafile :	<u>VW032531.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1125SBS01	Dichlorodifluoromethane	20	22.6	ug/Kg	113			64	136	
	Chloromethane	20	22.3	ug/Kg	112			73	129	
	Vinyl chloride	20	22.9	ug/Kg	115			73	133	
	Bromomethane	20	21.8	ug/Kg	109			77	137	
	Chloroethane	20	21.9	ug/Kg	110			69	130	
	Trichlorofluoromethane	20	19.0	ug/Kg	95			69	134	
	1,1,2-Trichlorotrifluoroethane	20	23.4	ug/Kg	117			81	123	
	1,1-Dichloroethene	20	22.9	ug/Kg	115			79	121	
	Acetone	100	110	ug/Kg	110			60	174	
	Carbon disulfide	20	21.2	ug/Kg	106			73	125	
	Methyl tert-butyl Ether	20	20.6	ug/Kg	103			77	129	
	Methyl Acetate	20	21.8	ug/Kg	109			51	140	
	Methylene Chloride	20	23.9	ug/Kg	119			54	158	
	trans-1,2-Dichloroethene	20	22.1	ug/Kg	111			80	123	
	1,1-Dichloroethane	20	23.4	ug/Kg	117			82	123	
	Cyclohexane	20	22.1	ug/Kg	111			76	122	
	2-Butanone	100	100	ug/Kg	100			69	131	
	Carbon Tetrachloride	20	23.9	ug/Kg	119			76	129	
	cis-1,2-Dichloroethene	20	22.2	ug/Kg	111			82	123	
	Bromochloromethane	20	22.5	ug/Kg	113			80	127	
	Chloroform	20	23.4	ug/Kg	117			82	125	
	1,1,1-Trichloroethane	20	22.9	ug/Kg	115			80	126	
	Methylcyclohexane	20	22.4	ug/Kg	112			77	123	
	Benzene	20	23.6	ug/Kg	118			84	121	
	1,2-Dichloroethane	20	24.1	ug/Kg	121			81	126	
	Trichloroethene	20	22.5	ug/Kg	113			83	122	
	1,2-Dichloropropane	20	24.4	ug/Kg	122			83	122	
	Bromodichloromethane	20	23.6	ug/Kg	118			82	123	
	4-Methyl-2-Pentanone	100	110	ug/Kg	110			70	135	
	Toluene	20	23.9	ug/Kg	119			83	122	
	t-1,3-Dichloropropene	20	21.7	ug/Kg	109			78	124	
	cis-1,3-Dichloropropene	20	22.4	ug/Kg	112			81	122	
	1,1,2-Trichloroethane	20	23.1	ug/Kg	116			82	125	
	2-Hexanone	100	110	ug/Kg	110			66	138	
	Dibromochloromethane	20	22.4	ug/Kg	112			79	125	
	1,2-Dibromoethane	20	23.2	ug/Kg	116			80	125	
	Tetrachloroethene	20	22.9	ug/Kg	115			83	125	
	Chlorobenzene	20	22.7	ug/Kg	114			84	122	
	Ethyl Benzene	20	22.1	ug/Kg	111			82	124	
	m/p-Xylenes	40	46.4	ug/Kg	116			83	124	
	o-Xylene	20	21.5	ug/Kg	108			83	123	
	Styrene	20	22.4	ug/Kg	112			82	124	
	Bromoform	20	21.0	ug/Kg	105			75	127	
	Isopropylbenzene	20	20.4	ug/Kg	102			82	124	
	1,1,2,2-Tetrachloroethane	20	21.2	ug/Kg	106			77	127	
	1,3-Dichlorobenzene	20	22.2	ug/Kg	111			83	122	
	1,4-Dichlorobenzene	20	21.3	ug/Kg	106			84	121	
	1,2-Dichlorobenzene	20	20.8	ug/Kg	104			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3719</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Remington & Vernick</u>	Datafile :	<u>VW032531.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1125SBS01	1,2-Dibromo-3-Chloropropane	20	18.4	ug/Kg	92			66	134	
	1,2,4-Trichlorobenzene	20	19.0	ug/Kg	95			78	127	
	1,2,3-Trichlorobenzene	20	19.6	ug/Kg	98			70	137	