

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3719	Analytical Method:	SW8260D
Client:	Remington & Vernick	Datafile :	VW032532.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1125SBSD01	Dichlorodifluoromethane	20	21.5	ug/Kg	108	5		64	136	20
	Chloromethane	20	21.7	ug/Kg	109	3		73	129	15
	Vinyl chloride	20	22.0	ug/Kg	110	4		73	133	15
	Bromomethane	20	21.7	ug/Kg	109	0		77	137	15
	Chloroethane	20	22.7	ug/Kg	114	4		69	130	15
	Trichlorofluoromethane	20	21.0	ug/Kg	105	10		69	134	27
	1,1,2-Trichlorotrifluoroethane	20	22.8	ug/Kg	114	3		81	123	20
	1,1-Dichloroethene	20	22.0	ug/Kg	110	4		79	121	16
	Acetone	100	110	ug/Kg	110	0		60	174	28
	Carbon disulfide	20	21.0	ug/Kg	105	1		73	125	15
	Methyl tert-butyl Ether	20	21.3	ug/Kg	106	3		77	129	20
	Methyl Acetate	20	22.7	ug/Kg	114	4		51	140	30
	Methylene Chloride	20	22.0	ug/Kg	110	8		54	158	20
	trans-1,2-Dichloroethene	20	21.6	ug/Kg	108	3		80	123	15
	1,1-Dichloroethane	20	22.6	ug/Kg	113	3		82	123	15
	Cyclohexane	20	21.6	ug/Kg	108	3		76	122	15
	2-Butanone	100	110	ug/Kg	110	10		69	131	29
	Carbon Tetrachloride	20	22.0	ug/Kg	110	8		76	129	15
	cis-1,2-Dichloroethene	20	22.1	ug/Kg	111	0		82	123	15
	Bromochloromethane	20	22.7	ug/Kg	114	1		80	127	15
	Chloroform	20	23.6	ug/Kg	118	1		82	125	15
	1,1,1-Trichloroethane	20	22.8	ug/Kg	114	1		80	126	15
	Methylcyclohexane	20	20.3	ug/Kg	102	9		77	123	15
	Benzene	20	21.8	ug/Kg	109	8		84	121	15
	1,2-Dichloroethane	20	22.6	ug/Kg	113	7		81	126	15
	Trichloroethene	20	21.3	ug/Kg	106	6		83	122	15
	1,2-Dichloropropane	20	22.7	ug/Kg	114	7		83	122	15
	Bromodichloromethane	20	22.1	ug/Kg	111	6		82	123	15
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	0		70	135	26
	Toluene	20	21.9	ug/Kg	110	8		83	122	15
	t-1,3-Dichloropropene	20	21.0	ug/Kg	105	4		78	124	15
	cis-1,3-Dichloropropene	20	20.9	ug/Kg	104	7		81	122	15
	1,1,2-Trichloroethane	20	22.1	ug/Kg	111	4		82	125	15
	2-Hexanone	100	110	ug/Kg	110	0		66	138	27
	Dibromochloromethane	20	21.3	ug/Kg	106	6		79	125	15
	1,2-Dibromoethane	20	21.8	ug/Kg	109	6		80	125	16
	Tetrachloroethene	20	21.3	ug/Kg	106	8		83	125	15
	Chlorobenzene	20	20.6	ug/Kg	103	10		84	122	15
	Ethyl Benzene	20	20.6	ug/Kg	103	7		82	124	15
	m/p-Xylenes	40	42.2	ug/Kg	106	9		83	124	15
	o-Xylene	20	20.1	ug/Kg	101	7		83	123	15
	Styrene	20	21.5	ug/Kg	108	4		82	124	15
	Bromoform	20	19.6	ug/Kg	98	7		75	127	16
	Isopropylbenzene	20	20.0	ug/Kg	100	2		82	124	15
	1,1,2,2-Tetrachloroethane	20	21.8	ug/Kg	109	3		77	127	21
	1,3-Dichlorobenzene	20	22.0	ug/Kg	110	1		83	122	15
	1,4-Dichlorobenzene	20	20.5	ug/Kg	103	3		84	121	15
	1,2-Dichlorobenzene	20	21.0	ug/Kg	105	1		83	124	15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3719</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Remington & Vernick</u>	Datafile :	<u>VW032532.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW1125SBSD01	1,2-Dibromo-3-Chloropropane	20	20.3	ug/Kg	102	10		66	134	20
	1,2,4-Trichlorobenzene	20	21.0	ug/Kg	105	10		78	127	20
	1,2,3-Trichlorobenzene	20	21.7	ug/Kg	109	11		70	137	16