

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI01
 Lab Code: ACE SDG NO.: Q3719
 Lab File ID: VW032529.D Date Analyzed: 11/25/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:00
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	353910	7.96	659809	8.85	585550	11.63
UPPER LIMIT	707820	8.459	1319620	9.349	1171100	12.129
LOWER LIMIT	176955	7.459	329905	8.349	292775	11.129
EPA SAMPLE NO.						
SB-1125-10A	210353	7.96	491657	8.85	525721	11.63
SB-1125-10B	227021	7.97	528970	8.85	565806	11.64
VW1125SBL01	312940	7.96	616508	8.85	570663	11.63
VW1125SBS01	347685	7.95	620529	8.85	580431	11.63
VW1125SBSD01	373267	7.97	705017	8.85	657789	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.