

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3719</u>
Lab File ID:	<u>VW032536.D</u>	Date Analyzed:	<u>11/26/2025</u>
Instrument ID:	<u>MSVOA_W</u>	Time Analyzed:	<u>12:05</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	353885	7.96	632151	8.85	586351	11.63
UPPER LIMIT	707770	8.459	1264300	9.349	1172700	12.129
LOWER LIMIT	176943	7.459	316076	8.349	293176	11.129
EPA SAMPLE NO.						
SB-1125-11A	193034	7.97	464802	8.86	514695	11.64
SB-1125-11B	183438	7.96	469951	8.85	532584	11.63
SB-1125-12A	186816	7.96	449736	8.86	514009	11.64
SB-1125-12B	186663	7.96	470881	8.85	539261	11.64
SB-1125-14A	200349	7.96	492902	8.85	550426	11.63
SB-1125-14B	191327	7.97	462395	8.85	513156	11.64
SB-1125-15A	176099 *	7.97	446493	8.85	510140	11.64
SB-1125-15B	181206	7.96	452149	8.86	513533	11.63
SB-1125-16A	179732	7.96	445052	8.85	515760	11.64
SB-1125-16B	189608	7.97	463920	8.85	526803	11.64
SB-1125-17B	212464	7.96	543417	8.85	642461	11.63
VW1126SBL01	322545	7.96	615508	8.85	556062	11.63
VW1126SBS01	337562	7.96	629179	8.85	578751	11.63

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.