

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI01
 Lab Code: ACE SDG NO.: Q3719
 Lab File ID: VY023829.D Date Analyzed: 12/01/2025
 Instrument ID: MSVOA_Y Time Analyzed: 07:55
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	565841	7.71	855261	8.62	736930	11.42
UPPER LIMIT	1131680	8.213	1710520	9.122	1473860	11.92
LOWER LIMIT	282921	7.213	427631	8.122	368465	10.92
EPA SAMPLE NO.						
SB-1125-15ARE	60762 *	7.71	124548 *	8.62	135409 *	11.42
SB-1125-17A	710245	7.71	1310532	8.62	1122921	11.42
VY1201SBL01	719047	7.71	1221329	8.62	995220	11.42
VY1201SBS01	535284	7.71	821167	8.62	693280	11.42

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.