

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3719</u>
Lab File ID:	<u>VW032536.D</u>	Date Analyzed:	<u>11/26/2025</u>
Instrument ID:	<u>MSVOA_W</u>	Time Analyzed:	<u>12:05</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	285431	13.555				
UPPER LIMIT	570862	14.055				
LOWER LIMIT	142716	13.055				
EPA SAMPLE NO.						
SB-1125-11A	246255	13.56				
SB-1125-11B	255138	13.56				
SB-1125-12A	231134	13.56				
SB-1125-12B	265861	13.56				
SB-1125-14A	245674	13.56				
SB-1125-14B	246111	13.56				
SB-1125-15A	256049	13.56				
SB-1125-15B	252269	13.56				
SB-1125-16A	254063	13.56				
SB-1125-16B	256014	13.56				
SB-1125-17B	328212	13.56				
VW1126SBL01	253580	13.56				
VW1126SBS01	285723	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.