

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3719

Lab File ID: VW032405.D

BFB Injection Date: 11/10/2025

Instrument ID: MSVOA_W

BFB Injection Time: 14:20

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.7
75	30.0 - 60.0% of mass 95	49.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 100.0% of mass 95	67.1
175	5.0 - 9.0% of mass 174	5.1 (7.6) 1
176	95.0 - 101.0% of mass 174	66.1 (98.5) 1
177	5.0 - 9.0% of mass 176	4.1 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VW032406.D	11/10/2025	15:11
VSTDIC010	VSTDIC010	VW032407.D	11/10/2025	15:33
VSTDIC020	VSTDIC020	VW032408.D	11/10/2025	15:55
VSTDIC050	VSTDIC050	VW032409.D	11/10/2025	16:17
VSTDIC100	VSTDIC100	VW032410.D	11/10/2025	16:39
VSTDIC150	VSTDIC150	VW032411.D	11/10/2025	17:00