

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3719

Lab File ID: VY023828.D

BFB Injection Date: 12/01/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 07:24

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.9
75	30.0 - 60.0% of mass 95	50.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.8 (1) 1
174	50.0 - 100.0% of mass 95	82.5
175	5.0 - 9.0% of mass 174	5.8 (7) 1
176	95.0 - 101.0% of mass 174	79.4 (96.3) 1
177	5.0 - 9.0% of mass 176	5.2 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023829.D	12/01/2025	07:55
VY1201SBL01	VY1201SBL01	VY023830.D	12/01/2025	09:13
VY1201SBS01	VY1201SBS01	VY023831.D	12/01/2025	09:37
SB-1125-15ARE	Q3719-09RE	VY023833.D	12/01/2025	10:38
SB-1125-17A	Q3719-13	VY023834.D	12/01/2025	11:25