

Initial and Continuing Calibration Verification

Client: Roman E&G Corp

SDG No.: Q3725

Project: AQUA Woolwich Pump Station 22-531

RunNo.: LB138051

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.098	0.099	99	90-110	11/26/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	11/26/2025
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	11/26/2025
Sample ID: Cyanide	CCV3	mg/L	0.24	0.25	96	90-110	11/26/2025
Sample ID: Cyanide	CCV4	mg/L	0.24	0.25	96	90-110	11/26/2025

Initial and Continuing Calibration Verification

Client: Roman E&G Corp

SDG No.: Q3725

Project: AQUA Woolwich Pump Station 22-531

RunNo.: LB138051

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------