

Continuing Calibration Report for SequenceID : FE120125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE057090.D

Aliphatic C9-C12	11796065.000	60.000	3.201	6.825	196601.083	200655.681	2.021
Aliphatic C12-C16	7910112.000	40.000	6.826	10.270	197752.800	203177.346	2.670
Aliphatic C16-C21	11460181.000	60.000	10.271	13.644	191003.017	198315.768	3.687
Aliphatic C21-C28	14258034.000	80.000	13.645	17.313	178225.425	198747.045	10.325
Aliphatic C28-C40	19087397.000	120.000	17.314	22.242	159061.642	199899.890	20.429
Aliphatic EPH	64511789.000	360.000	3.201	22.242	179199.414	199869.809	10.342